

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012920\
 Data File : VU036593.D
 Acq On : 29 Jan 2020 15:24
 Operator : JC/MD
 Sample : L1271-10 20X
 Misc : 5.10g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAT01

Quant Time: Jan 30 04:09:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 30 02:48:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	321371	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	302584	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	157865	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92683	38.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.00%
7) Chloroethane-d5	1.93	69	81448	39.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.88%
11) 1,1-Dichloroethene-d2	2.59	63	123145	31.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.28%
21) 2-Butanone-d5	4.67	46	128157	89.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.70%
24) Chloroform-d	5.11	84	161314	40.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.86%
26) 1,2-Dichloroethane-d4	5.74	65	101678	41.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.78%
32) Benzene-d6	5.76	84	365117	42.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.80%
36) 1,2-Dichloropropane-d6	6.73	67	112981	42.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.20%
41) Toluene-d8	7.93	98	354053	43.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.48%
43) trans-1,3-Dichloropropene-	8.21	79	54903	43.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.04%
47) 2-Hexanone-d5	8.66	63	116106	94.40	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.40%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	164152	42.09	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.18%
64) 1,2-Dichlorobenzene-d4	12.21	152	143263	42.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012920\
Data File : VU036593.D
Acq On : 29 Jan 2020 15:24
Operator : JC/MD
Sample : L1271-10 20X
Misc : 5.10µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAT01

Quant Time: Jan 30 04:09:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 30 02:48:11 2020
Response via : Initial Calibration

