

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
 Data File : VU042217.D
 Acq On : 29 Jan 2021 10:07
 Operator : SY/MD
 Sample : VU0129WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0129WBL01

Quant Time: Jan 30 00:46:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jan 29 09:59:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.125	96	13134	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	5167	1.01	ug/l	0.00
Spiked Amount 1.000			Recovery	=	101.00%	
68) 1,2-Dichlorobenzene-d4	12.205	152	5117	0.99	ug/l	0.00
Spiked Amount 1.000			Recovery	=	99.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
Data File : VU042217.D
Acq On : 29 Jan 2021 10:07
Operator : SY/MD
Sample : VU0129WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0129WBL01

Quant Time: Jan 30 00:46:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jan 29 09:59:59 2021
Response via : Initial Calibration

