

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
 Data File : VU042220.D
 Acq On : 29 Jan 2021 11:40
 Operator : SY/MD
 Sample : M1126-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jan 30 00:48:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jan 29 09:59:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.124	96	13306	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	4999	0.97	ug/l	0.00
Spiked Amount 1.000			Recovery	=	97.00%	
68) 1,2-Dichlorobenzene-d4	12.205	152	4690	0.89	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.00%	
Target Compounds						
27) Chloroform	5.102	83	26714	4.14	ug/l	Qvalue 98
44) Bromodichloromethane	7.118	83	47712	9.27	ug/l	100
55) Dibromochloromethane	8.822	129	30659	8.85	ug/l	99
67) Bromoform	10.301	173	20034	9.19	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
Data File : VU042220.D
Acq On : 29 Jan 2021 11:40
Operator : SY/MD
Sample : M1126-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jan 30 00:48:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jan 29 09:59:59 2021
Response via : Initial Calibration

