

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
 Data File : VU042222.D
 Acq On : 29 Jan 2021 13:29
 Operator : SY/MD
 Sample : M1126-10DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WSDL

Quant Time: Jan 30 01:19:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jan 29 09:59:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.125	96	12788	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	4909	0.99	ug/l	0.00
Spiked Amount 1.000			Recovery	=	99.00%	
68) 1,2-Dichlorobenzene-d4	12.205	152	4958	0.98	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.00%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	13598	4.19	ug/l	99
24) tert-Butyl Alcohol	3.208	59	1587	8.97	ug/l #	72
25) Methyl tert-Butyl Ether	3.379	73	26827	3.10	ug/l	97
31) Isopropyl Ether	4.012	45	91823	7.88	ug/l	99
32) Ethyl-t-butyl ether	4.523	59	74588	7.12	ug/l	99
33) Tert-Amyl methyl ether	5.961	73	33747	3.45	ug/l	99
73) n-propylbenzene	10.915	120	31505	6.90	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012921\
Data File : VU042222.D
Acq On : 29 Jan 2021 13:29
Operator : SY/MD
Sample : M1126-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Quant Time: Jan 30 01:19:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jan 29 09:59:59 2021
Response via : Initial Calibration

