

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012924\
 Data File : VU057714.D
 Acq On : 29 Jan 2024 14:05
 Operator : MD/SY
 Sample : P1230-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BG0

Quant Time: Jan 30 00:33:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 30 00:30:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	67104	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61659	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28344	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.911	69	17037	5.499	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.000%
11) 1,1-Dichloroethene-d2	2.563	65	2900	1.668	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	33.400%#
20) 2-Butanone-d5	4.634	46	41250	59.141	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.280%
24) Chloroform-d	5.055	84	38864	5.179	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.695	65	19656	5.764	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.200%
32) Benzene-d6	5.721	84	79295	5.483	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.600%
36) 1,2-Dichloropropane-d6	6.685	67	24309	5.572	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.400%
41) Toluene-d8	7.898	98	2448	0.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	3.600%#
43) trans-1,3-Dichloroprop...	8.181	79	1254	0.721	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	14.400%#
46) 2-Hexanone-d5	8.631	63	31684	57.116	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.240%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	15643	5.350	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	22806	5.733	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.600%
Target Compounds						Qvalue
38) Bromodichloromethane	7.103	83	1587	0.223	ug/L #	96
49) Dibromochloromethane	8.804	129	2647	0.594	ug/L	99
59) Bromoform	10.287	173	1003	0.436	ug/L #	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012924\
Data File : VU057714.D
Acq On : 29 Jan 2024 14:05
Operator : MD/SY
Sample : P1230-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BG0

Quant Time: Jan 30 00:33:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 30 00:30:52 2024
Response via : Initial Calibration









#38
Bromodichloromethane
Concen: 0.223 ug/L
RT: 7.103 min Scan# 11
Delta R.T. 0.003 min
Lab File: VU057714.D
Acq: 29 Jan 2024 14:05




Instrument :
MSVOA_U
ClientSampleId :
C0BG0


Tgt Ion: 83 Resp: 1587
Ion Ratio Lower Upper
83 100
85 62.6 44.7 83.1
127 0.0 6.4 9.6#



#49
Dibromochloromethane
Concen: 0.594 ug/L
RT: 8.804 min Scan# 2337
Delta R.T. 0.000 min
Lab File: VU057714.D
Acq: 29 Jan 2024 14:05


Tgt Ion: 129 Resp: 2647
Ion Ratio Lower Upper
129 100
127 78.6 54.5 101.1





#59
 Bromoform
 Concen: 0.436 ug/L
 RT: 10.287 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VU057714.D
 Acq: 29 Jan 2024 14:05

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BG0



Tgt Ion	Ratio	Lower	Upper
173	100		
175	51.5	37.5	56.3
254	0.0	7.2	10.8

