

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012925\
 Data File : VU063088.D
 Acq On : 29 Jan 2025 12:49
 Operator : MD/SY
 Sample : VIBLK006
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK006

Quant Time: Jan 30 07:35:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 30 07:33:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	97999	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	92978	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	53271	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	10570	3.000	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.000%
7) Chloroethane-d5	1.904	69	10882	4.068	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.400%
11) 1,1-Dichloroethene-d2	2.557	65	7936	3.767	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.400%
20) 2-Butanone-d5	4.605	46	42990	42.046	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.100%
24) Chloroform-d	5.049	84	54734	3.911	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
26) 1,2-Dichloroethane-d4	5.689	65	31387	3.929	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
32) Benzene-d6	5.714	84	80119	3.771	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.400%
36) 1,2-Dichloropropane-d6	6.679	67	20198	3.733	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.600%
41) Toluene-d8	7.888	98	78248	3.725	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.400%
43) trans-1,3-Dichloroprop...	8.174	79	12678	3.785	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	8.621	63	38697	46.307	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.620%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	22223	4.326	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.600%
66) 1,2-Dichlorobenzene-d4	12.184	152	40621	4.260	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012925\
Data File : VU063088.D
Acq On : 29 Jan 2025 12:49
Operator : MD/SY
Sample : VIBLK006
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK006

Quant Time: Jan 30 07:35:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 30 07:33:13 2025
Response via : Initial Calibration

