

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU013020\
 Data File : VU036610.D
 Acq On : 30 Jan 2020 11:39
 Operator : JC/MD
 Sample : VU0130MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK50

Quant Time: Jan 30 13:12:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 30 12:57:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	299841	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	279989	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	130850	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92660	41.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.52%
7) Chloroethane-d5	1.93	69	83314	43.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.48%
11) 1,1-Dichloroethene-d2	2.59	63	121582	33.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.96%
21) 2-Butanone-d5	4.68	46	143473	107.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.63%
24) Chloroform-d	5.11	84	159950	43.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.98%
26) 1,2-Dichloroethane-d4	5.75	65	107584	46.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.86%
32) Benzene-d6	5.77	84	369216	46.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.76%
36) 1,2-Dichloropropane-d6	6.73	67	119455	48.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.34%
41) Toluene-d8	7.93	98	352507	46.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.04%
43) trans-1,3-Dichloropropene-	8.21	79	57236	49.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.06%
47) 2-Hexanone-d5	8.67	63	131412	115.46	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.46%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	175511	48.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.28%
64) 1,2-Dichlorobenzene-d4	12.21	152	131072	46.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.36%

Target Compounds					Ovalue
13) Acetone	2.66	43	2512	1.947 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU013020\
Data File : VU036610.D
Acq On : 30 Jan 2020 11:39
Operator : JC/MD
Sample : VU0130MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK50

Quant Time: Jan 30 13:12:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 30 12:57:35 2020
Response via : Initial Calibration

