

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU013020\
 Data File : VU036615.D
 Acq On : 30 Jan 2020 13:37
 Operator : JC/MD
 Sample : L1301-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFM07

Quant Time: Jan 30 15:39:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 30 12:57:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	299388	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	278330	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	131721	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	93091	41.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.02%
7) Chloroethane-d5	1.93	69	83822	43.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.14%
11) 1,1-Dichloroethene-d2	2.59	63	121043	33.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.76%
21) 2-Butanone-d5	4.68	46	142537	107.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.09%
24) Chloroform-d	5.11	84	164050	44.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.36%
26) 1,2-Dichloroethane-d4	5.74	65	105666	46.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.34%
32) Benzene-d6	5.77	84	367100	46.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.78%
36) 1,2-Dichloropropane-d6	6.73	67	117040	47.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.94%
41) Toluene-d8	7.93	98	349764	46.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.88%
43) trans-1,3-Dichloropropene-	8.21	79	54073	46.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.20%
47) 2-Hexanone-d5	8.67	63	123383	109.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.05%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	168930	47.09	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.18%
64) 1,2-Dichlorobenzene-d4	12.21	152	132826	46.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.98%

Target Compounds

					Ovalue
13) Acetone	2.66	43	3223	2.502	ug/L 100
33) Benzene	5.82	78	94008	11.297	ug/L 100
42) Toluene	8.00	91	33175	3.668	ug/L 99
53) m,p-Xylene	9.72	106	5535	1.432	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU013020\
Data File : VU036615.D
Acq On : 30 Jan 2020 13:37
Operator : JC/MD
Sample : L1301-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFM07

Quant Time: Jan 30 15:39:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 30 12:57:35 2020
Response via : Initial Calibration

