

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU013020\
 Data File : VU036637.D
 Acq On : 30 Jan 2020 23:20
 Operator : JC/MD
 Sample : L1324-04 10X
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAT61

Quant Time: Jan 31 06:15:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 31 04:01:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	287366	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	282854	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	151140	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	69337	32.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.42%
7) Chloroethane-d5	1.89	69	20312	11.00	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	22.00%#
11) 1,1-Dichloroethene-d2	2.57	63	95731	27.51	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.02%#
21) 2-Butanone-d5	4.68	46	98468	77.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.07%
24) Chloroform-d	5.11	84	135940	38.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.14%
26) 1,2-Dichloroethane-d4	5.74	65	92087	41.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.84%
32) Benzene-d6	5.76	84	323134	40.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.22%
36) 1,2-Dichloropropane-d6	6.73	67	104292	42.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.12%
41) Toluene-d8	7.93	98	311591	40.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.42%
43) trans-1,3-Dichloropropene-	8.21	79	47494	40.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.56%
47) 2-Hexanone-d5	8.67	63	128345	111.63	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.63%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	162151	44.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.96%
64) 1,2-Dichlorobenzene-d4	12.21	152	138749	42.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.56%

Target Compounds

					Ovalue
13) Acetone	2.66	43	2515	2.034	ug/L 98
33) Benzene	5.81	78	12468	1.474	ug/L 100
42) Toluene	8.00	91	732872	79.729	ug/L 98
46) Tetrachloroethene	8.58	164	112726	63.748	ug/L 97
52) Ethylbenzene	9.60	91	197989	19.662	ug/L 99
53) m,p-Xylene	9.72	106	581340	148.022	ug/L 98
54) o-xylene	10.13	106	263181	67.181	ug/L 98
55) Styrene	10.14	104	691922	108.838	ug/L 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU013020\
Data File : VU036637.D
Acq On : 30 Jan 2020 23:20
Operator : JC/MD
Sample : L1324-04 10X
Misc : 5.00uL/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAT61

Quant Time: Jan 31 06:15:52 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM012920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 31 04:01:38 2020
Response via : Initial Calibration

