

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU013023\  
 Data File : VU052862.D  
 Acq On : 30 Jan 2023 11:52  
 Operator : JC/MD  
 Sample : VU0130WBL01  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VU0130WBL01

Quant Time: Jan 31 07:41:27 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U012323DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Tue Jan 24 05:16:59 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| -----                       |        |      |          |       |         |          |
| Internal Standards          |        |      |          |       |         |          |
| 1) Fluorobenzene            | 6.112  | 96   | 33860    | 1.000 | ug/l    | # 0.00   |
|                             |        |      |          |       |         |          |
| System Monitoring Compounds |        |      |          |       |         |          |
| 57) 4-Bromofluorobenzene    | 10.632 | 95   | 8245     | 0.732 | ug/l    | 0.00     |
| Spiked Amount 1.000         |        |      | Recovery | =     | 73.000% |          |
| 68) 1,2-Dichlorobenzene-d4  | 12.189 | 152  | 9952     | 0.859 | ug/l    | 0.00     |
| Spiked Amount 1.000         |        |      | Recovery | =     | 86.000% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU013023\  
Data File : VU052862.D  
Acq On : 30 Jan 2023 11:52  
Operator : JC/MD  
Sample : VU0130WBL01  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0130WBL01

Quant Time: Jan 31 07:41:27 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U012323DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Tue Jan 24 05:16:59 2023  
Response via : Initial Calibration

