

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013023\
 Data File : VU052864.D
 Acq On : 30 Jan 2023 13:28
 Operator : JC/MD
 Sample : 01245-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-EDBCP-WS

Quant Time: Jan 31 07:43:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.111	96	32997	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	8750	0.797	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	10207	0.904	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.000%	
Target Compounds						
						Qvalue
56) 1,2-Dibromoethane	8.925	107	2840	0.466	ug/l	84
71) 1,2,3-Trichloropropane	10.822	75	9330	1.352	ug/l	88
86) 1,2-Dibromo-3-Chloropr...	12.995	75	568	0.385	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013023\
Data File : VU052864.D
Acq On : 30 Jan 2023 13:28
Operator : JC/MD
Sample : 01245-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-EDBCP-WS

Quant Time: Jan 31 07:43:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration

