

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013023\
 Data File : VU052865.D
 Acq On : 30 Jan 2023 16:10
 Operator : JC/MD
 Sample : 01245-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jan 31 07:44:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.108	96	33590	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	8602	0.770	ug/l	0.00
Spiked Amount 1.000			Recovery	=	77.000%	
68) 1,2-Dichlorobenzene-d4	12.188	152	9471	0.824	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
27) Chloroform	5.086	83	168811	10.695	ug/l	98
44) Bromodichloromethane	7.105	83	95723	7.706	ug/l	99
55) Dibromochloromethane	8.806	129	76848	9.023	ug/l	99
67) Bromoform	10.288	173	45774	9.108	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013023\
Data File : VU052865.D
Acq On : 30 Jan 2023 16:10
Operator : JC/MD
Sample : 01245-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jan 31 07:44:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration

