

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013024\
 Data File : VU057737.D
 Acq On : 30 Jan 2024 12:33
 Operator : MD/SY
 Sample : P1230-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BE9

Quant Time: Jan 30 22:36:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 30 22:31:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	67729	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61982	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	27380	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	18291	4.902	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.000%
7) Chloroethane-d5	1.911	69	15532	4.967	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.563	65	8647	4.928	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.600%
20) 2-Butanone-d5	4.637	46	36819	52.301	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.600%
24) Chloroform-d	5.055	84	37514	4.953	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
26) 1,2-Dichloroethane-d4	5.695	65	18076	5.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.000%
32) Benzene-d6	5.721	84	73681	5.068	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
36) 1,2-Dichloropropane-d6	6.682	67	22003	5.018	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%
41) Toluene-d8	7.891	98	66862	4.845	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
43) trans-1,3-Dichloroprop...	8.177	79	8050	4.606	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.200%
46) 2-Hexanone-d5	8.631	63	27791	49.838	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.680%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	14189	4.827	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	20809	5.416	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%
Target Compounds						Qvalue
25) Chloroform	5.081	83	44664	4.397	ug/L	98
38) Bromodichloromethane	7.100	83	8619	1.204	ug/L	100
49) Dibromochloromethane	8.804	129	1248	0.279	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013024\
Data File : VU057737.D
Acq On : 30 Jan 2024 12:33
Operator : MD/SY
Sample : P1230-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BE9

Quant Time: Jan 30 22:36:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 30 22:31:02 2024
Response via : Initial Calibration

