

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013024\
 Data File : VU057753.D
 Acq On : 30 Jan 2024 19:00
 Operator : MD/SY
 Sample : P1239-13DL 400X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FC3DL

Quant Time: Jan 30 22:39:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 30 22:31:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	63537	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	57935	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	26153	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	16466	4.704	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.000%	
7) Chloroethane-d5	1.911	69	14626	4.986	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.800%	
11) 1,1-Dichloroethene-d2	2.563	65	7509	4.562	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.200%	
20) 2-Butanone-d5	4.627	46	34393	52.079	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.160%	
24) Chloroform-d	5.055	84	32989	4.643	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	5.695	65	17068	5.286	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.800%	
32) Benzene-d6	5.721	84	66398	4.886	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.800%	
36) 1,2-Dichloropropane-d6	6.685	67	20570	5.018	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.400%	
41) Toluene-d8	7.894	98	59875	4.641	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.800%	
43) trans-1,3-Dichloroprop...	8.177	79	7111	4.353	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.000%	
46) 2-Hexanone-d5	8.631	63	25458	48.843	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.680%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	12897	4.694	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	18791	5.120	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	
Target Compounds						Qvalue
5) Vinyl chloride	1.602	62	1922	0.335	ug/L	84
22) cis-1,2-Dichloroethene	4.660	96	53270	10.188	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013024\
Data File : VU057753.D
Acq On : 30 Jan 2024 19:00
Operator : MD/SY
Sample : P1239-13DL 400X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FC3DL

Quant Time: Jan 30 22:39:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 30 22:31:02 2024
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 0.335 ug/L
 RT: 1.602 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VU057753.D
 Acq: 30 Jan 2024 19:00

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FC3DL

Tgt Ion: 62 Resp: 1922
 Ion Ratio Lower Upper
 62 100
 64 41.5 22.8 42.4



#22
 cis-1,2-Dichloroethene
 Concen: 10.188 ug/L
 RT: 4.660 min Scan# 1048
 Delta R.T. -0.000 min
 Lab File: VU057753.D
 Acq: 30 Jan 2024 19:00

Tgt Ion: 96 Resp: 53270
 Ion Ratio Lower Upper
 96 100
 61 134.7 92.7 172.1
 98 65.2 44.9 83.5

