

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU013120\
 Data File : VU036664.D
 Acq On : 01 Feb 2020 01:09
 Operator : JC/MD
 Sample : L1323-18 20X
 Misc : 4.99g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAT56

Quant Time: Feb 15 01:36:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM013120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 01 00:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	273045	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	257261	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	117495	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	117385	47.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.32%
7) Chloroethane-d5	1.92	69	99308	49.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.10%
11) 1,1-Dichloroethene-d2	2.59	63	141522	36.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.66%
21) 2-Butanone-d5	4.67	46	131135	85.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.57%
24) Chloroform-d	5.11	84	165923	44.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.98%
26) 1,2-Dichloroethane-d4	5.75	65	121816	50.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.84%
32) Benzene-d6	5.77	84	387739	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.56%
36) 1,2-Dichloropropane-d6	6.73	67	128657	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.54%
41) Toluene-d8	7.93	98	349989	48.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.40%
43) trans-1,3-Dichloropropene-	8.21	79	58912	47.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.62%
47) 2-Hexanone-d5	8.67	63	152012	106.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.54%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	176011	48.38	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.76%
64) 1,2-Dichlorobenzene-d4	12.21	152	123068	49.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds

					Ovalue
25) Chloroform	5.13	83	10735	3.037 ug/L	98
33) Benzene	5.82	78	10717	1.307 ug/L	100
46) Tetrachloroethene	8.58	164	2139	1.458 ug/L	97
52) Ethylbenzene	9.60	91	31371	3.253 ug/L	99
53) m,p-Xylene	9.72	106	77946	21.914 ug/L	99
54) o-xylene	10.12	106	38327	10.697 ug/L	96
55) Styrene	10.14	104	108530	17.819 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU013120\
Data File : VU036664.D
Acq On : 01 Feb 2020 01:09
Operator : JC/MD
Sample : L1323-18 20X
Misc : 4.99g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAT56

Quant Time: Feb 15 01:36:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM013120WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 01 00:19:17 2020
Response via : Initial Calibration

