

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU020119\
 Data File : VU029336.D
 Acq On : 01 Feb 2019 17:07
 Operator : JC/SP
 Sample : K1287-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Feb 02 00:19:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR020119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 01 23:50:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	61310	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	63049	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	29558	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	24705	5.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.00%
7) Chloroethane-d5	1.68	69	24902	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.60%
11) 1,1-Dichloroethene-d2	2.27	63	50755	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.60%
20) 2-Butanone-d5	4.19	46	78610	50.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.28%
24) Chloroform-d	4.65	84	55818	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	5.31	65	35399	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.34	84	93428	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.33	67	29757	5.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.20%
41) Toluene-d8	7.56	98	80305	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.85	79	13596	4.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.20%
46) 2-Hexanone-d5	8.31	63	59037	49.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.48%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	24484	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	33030	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU020119\
Data File : VU029336.D
Acq On : 01 Feb 2019 17:07
Operator : JC/SP
Sample : K1287-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Feb 02 00:19:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR020119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 01 23:50:10 2019
Response via : Initial Calibration

