

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU020119\
 Data File : VU029333.D
 Acq On : 01 Feb 2019 15:08
 Operator : JC/SP
 Sample : VU0201WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK48

Quant Time: Feb 02 00:07:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR020119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 01 23:50:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	78728	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	79718	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	37477	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	27611	4.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.00%
7) Chloroethane-d5	1.68	69	25988	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.20%
11) 1,1-Dichloroethene-d2	2.27	63	52477	3.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.60%
20) 2-Butanone-d5	4.20	46	81136	40.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.40%
24) Chloroform-d	4.65	84	58302	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
26) 1,2-Dichloroethane-d4	5.31	65	37642	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
32) Benzene-d6	5.34	84	101447	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
36) 1,2-Dichloropropane-d6	6.33	67	32164	4.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.60%
41) Toluene-d8	7.56	98	87333	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
43) trans-1,3-Dichloropropene-	7.85	79	15020	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.31	63	62748	41.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.78%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	26337	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	34730	4.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU020119\
Data File : VU029333.D
Acq On : 01 Feb 2019 15:08
Operator : JC/SP
Sample : VU0201WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK48

Quant Time: Feb 02 00:07:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR020119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 01 23:50:10 2019
Response via : Initial Calibration

