

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020124\
 Data File : VU057780.D
 Acq On : 01 Feb 2024 16:07
 Operator : MD/SY
 Sample : P1174-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Feb 02 00:35:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U013124DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 02 00:27:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	41617	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	14245	0.904	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	12215	0.861	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.042	84	6571	0.532	ug/l	93
27) Chloroform	5.081	83	205971	9.293	ug/l	96
44) Bromodichloromethane	7.103	83	27707	1.780	ug/l #	95
55) Dibromochloromethane	8.804	129	61313	6.467	ug/l	96
67) Bromoform	10.286	173	44844	9.189	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020124\
Data File : VU057780.D
Acq On : 01 Feb 2024 16:07
Operator : MD/SY
Sample : P1174-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Feb 02 00:35:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U013124DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 02 00:27:55 2024
Response via : Initial Calibration

