

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020224\
 Data File : VU057789.D
 Acq On : 02 Feb 2024 12:08
 Operator : MD/SY
 Sample : VIBLK
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Feb 03 05:54:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U013124DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 02 00:27:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.110	96	41202	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.634	95	13806	0.885	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	12300	0.875	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
Target Compounds						
85) 1,2-Dichlorobenzene	12.209	146	11188	0.530	ug/l	97
87) 1,2,4-Trichlorobenzene	13.839	180	18894	1.499	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020224\
Data File : VU057789.D
Acq On : 02 Feb 2024 12:08
Operator : MD/SY
Sample : VIBLK
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Feb 03 05:54:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U013124DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 02 00:27:55 2024
Response via : Initial Calibration

