

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020321\
 Data File : VU042242.D
 Acq On : 03 Feb 2021 19:02
 Operator : SY/MD
 Sample : M1259-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E60

Quant Time: Feb 04 04:25:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 04 04:20:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	102406	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	94413	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	50753	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	19762	2.99	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.80%
7) Chloroethane-d5	1.919	69	18672	3.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.20%
11) 1,1-Dichloroethene-d2	2.575	63	36343	2.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.00%#
20) 2-Butanone-d5	4.645	46	127123	52.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.48%
24) Chloroform-d	5.073	84	56991	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.60%
26) 1,2-Dichloroethane-d4	5.713	65	38596	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.739	84	103171	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
36) 1,2-Dichloropropane-d6	6.700	67	36577	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.00%
41) Toluene-d8	7.906	98	96405	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.80%
43) trans-1,3-Dichloroprop...	8.189	79	15522	3.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.00%
46) 2-Hexanone-d5	8.645	63	91079	51.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.80%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	33294	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	43376	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%
Target Compounds						Qvalue
5) Vinyl chloride	1.607	62	51503	6.25	ug/L	96
13) Acetone	2.658	43	6733	4.19	ug/L	88
30) Cyclohexane	5.398	56	15115	1.13	ug/L	99
42) Toluene	7.980	91	9617	0.32	ug/L	97
52) Ethylbenzene	9.578	91	9840	0.29	ug/L	94
53) m,p-Xylene	9.700	106	18764	1.45	ug/L	96
54) o-Xylene	10.108	106	14227	1.14	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020321\
Data File : VU042242.D
Acq On : 03 Feb 2021 19:02
Operator : SY/MD
Sample : M1259-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E60

Quant Time: Feb 04 04:25:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 04 04:20:04 2021
Response via : Initial Calibration

