

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020321\
 Data File : VU042253.D
 Acq On : 03 Feb 2021 23:50
 Operator : SY/MD
 Sample : M1298-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C2TY1

Quant Time: Feb 04 05:36:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 04 05:34:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	99948	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	91245	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	48417	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	18227	2.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	56.60%
7) Chloroethane-d5	1.919	69	18125	3.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.80%
11) 1,1-Dichloroethene-d2	2.575	63	35389	2.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.80%#
20) 2-Butanone-d5	4.645	46	122101	51.90	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.80%
24) Chloroform-d	5.073	84	56609	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
26) 1,2-Dichloroethane-d4	5.713	65	37803	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	5.735	84	100266	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.20%
36) 1,2-Dichloropropane-d6	6.700	67	36291	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.906	98	93300	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.80%
43) trans-1,3-Dichloroprop...	8.189	79	14451	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	8.645	63	87875	51.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.62%
57) 1,1,2,2-Tetrachloroeth...	10.767	84	32673	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
64) 1,2-Dichlorobenzene-d4	12.201	152	41998	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020321\
Data File : VU042253.D
Acq On : 03 Feb 2021 23:50
Operator : SY/MD
Sample : M1298-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C2TY1

Quant Time: Feb 04 05:36:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 04 05:34:35 2021
Response via : Initial Calibration

