


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
Data File : VU042265.D
Acq On : 04 Feb 2021 14:40
Operator : SY/MD
Sample : M1323-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E67

Manual Integrations
APPROVED

MMDadoda
2/5/2021 9:40:38 AM

Quant Time: Feb 05 02:25:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 05 02:23:43 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
Data File : VU042265.D
Acq On : 04 Feb 2021 14:40
Operator : SY/MD
Sample : M1323-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E67

Manual Integrations
APPROVED

MMDadoda
2/5/2021 9:40:38 AM

Quant Time: Feb 08 02:48:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 05 02:23:43 2021
Response via : Initial Calibration



TIC: VU042265.D\data.ms

(13) Acetone (T)

2.662min (+ 0.009) 0.90 ug/L

response 1344

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.90	8.48
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
 Data File : VU042265.D
 Acq On : 04 Feb 2021 14:40
 Operator : SY/MD
 Sample : M1323-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E67

Manual Integrations
 APPROVED

MMDadoda
 2/5/2021 9:40:38 AM

Quant Time: Feb 05 02:25:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 05 02:23:43 2021
 Response via : Initial Calibration



TIC: VU042265.D\data.ms

(13) Acetone (T)

2.662min (+ 0.009) 1.40 ug/L m

response 2080

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.90	5.48
0.00	0.00	0.00
0.00	0.00	0.00

Handwritten: MMDadoda
 02/08/21

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
 Data File : VU042265.D
 Acq On : 04 Feb 2021 14:40
 Operator : SY/MD
 Sample : M1323-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E67

Manual Integrations
 APPROVED

MMDadoda
 2/5/2021 9:40:38 AM

Quant Time: Feb 05 02:25:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 05 02:23:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	95074	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	89045	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	46460	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	26706	4.35	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	87.00%		
7) Chloroethane-d5	1.919	69	23357	4.73	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	94.60%		
11) 1,1-Dichloroethene-d2	2.575	63	49971	3.70	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	74.00%		
20) 2-Butanone-d5	4.649	46	112418	50.23	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	100.46%		
24) Chloroform-d	5.073	84	60291	4.71	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	94.20%		
26) 1,2-Dichloroethane-d4	5.713	65	38233	4.99	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	99.80%		
32) Benzene-d6	5.735	84	113309	4.87	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	97.40%		
36) 1,2-Dichloropropane-d6	6.700	67	38236	5.10	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	102.00%		
41) Toluene-d8	7.906	98	105764	4.81	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	96.20%		
43) trans-1,3-Dichloroprop...	8.189	79	15173	3.99	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	79.80%		
46) 2-Hexanone-d5	8.645	63	80418	48.58	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	97.16%		
57) 1,1,2,2-Tetrachloroeth...	10.764	84	30922	4.81	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	96.20%		
64) 1,2-Dichlorobenzene-d4	12.201	152	42020	5.03	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	100.60%		
Target Compounds						
13) Acetone	2.662	43	2080m	1.40	ug/L	Qvalue
17) Methyl tert-butyl Ether	3.375	73	4399	0.24	ug/L #	87
22) cis-1,2-Dichloroethene	4.678	96	32619	4.46	ug/L	99
25) Chloroform	5.102	83	3827	0.29	ug/L	92
34) Trichloroethene	6.552	95	16904	2.39	ug/L	98
47) Tetrachloroethene	8.562	164	157285	26.77	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

7178
 02/08/21