

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
 Data File : VU042262.D
 Acq On : 04 Feb 2021 11:34
 Operator : SY/MD
 Sample : M1259-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Feb 05 02:24:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 05 02:23:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	91071	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	84349	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	43789	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	28815	4.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.00%
7) Chloroethane-d5	1.919	69	24666	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.20%
11) 1,1-Dichloroethene-d2	2.575	63	53603	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.80%
20) 2-Butanone-d5	4.642	46	89234	41.63	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.26%
24) Chloroform-d	5.073	84	63862	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	5.713	65	40793	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.20%
32) Benzene-d6	5.739	84	120464	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.40%
36) 1,2-Dichloropropane-d6	6.700	67	40141	5.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.00%
41) Toluene-d8	7.906	98	111899	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
43) trans-1,3-Dichloroprop...	8.189	79	16328	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
46) 2-Hexanone-d5	8.645	63	84876	54.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.26%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	31850	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	43358	5.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
Data File : VU042262.D
Acq On : 04 Feb 2021 11:34
Operator : SY/MD
Sample : M1259-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Feb 05 02:24:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 05 02:23:43 2021
Response via : Initial Calibration

