

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
 Data File : VU042263.D
 Acq On : 04 Feb 2021 13:52
 Operator : SY/MD
 Sample : M1323-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E75

Quant Time: Feb 05 02:24:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 05 02:23:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	114381	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	105841	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	56592	5.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	25210	3.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.40%
7) Chloroethane-d5	1.919	69	21582	3.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.60%
11) 1,1-Dichloroethene-d2	2.575	63	46537	2.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.20%#
20) 2-Butanone-d5	4.645	46	78616	29.20	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	58.40%
24) Chloroform-d	5.073	84	56478	3.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.40%
26) 1,2-Dichloroethane-d4	5.716	65	34706	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.20%
32) Benzene-d6	5.739	84	104566	3.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.60%
36) 1,2-Dichloropropane-d6	6.700	67	34800	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.00%
41) Toluene-d8	7.909	98	98241	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.20%
43) trans-1,3-Dichloroprop...	8.189	79	15015	3.32	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.40%
46) 2-Hexanone-d5	8.645	63	74805	38.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.04%
57) 1,1,2,2-Tetrachloroeth...	10.767	84	29291	3.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.60%
64) 1,2-Dichlorobenzene-d4	12.205	152	38951	3.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	76.60%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020421\
Data File : VU042263.D
Acq On : 04 Feb 2021 13:52
Operator : SY/MD
Sample : M1323-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E75

Quant Time: Feb 05 02:24:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 05 02:23:43 2021
Response via : Initial Calibration

