

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
 Data File : VU042278.D
 Acq On : 05 Feb 2021 13:11
 Operator : SY/MD
 Sample : VU0205WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK110

Quant Time: Feb 06 07:24:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 06 07:13:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	184364	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	164690	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	81241	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	68096	49.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.16%
7) Chloroethane-d5	1.916	69	50261	52.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.64%
11) 1,1-Dichloroethene-d2	2.571	63	114322	38.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.22%
21) 2-Butanone-d5	4.639	46	146517	102.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.07%
24) Chloroform-d	5.073	84	137616	49.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.78%
26) 1,2-Dichloroethane-d4	5.713	65	101629	51.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.10%
32) Benzene-d6	5.735	84	256423	53.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.18%
36) 1,2-Dichloropropane-d6	6.700	67	86406	54.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.12%
41) Toluene-d8	7.906	98	235303	51.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.82%
43) trans-1,3-Dichloroprop...	8.189	79	43014	51.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.54%
47) 2-Hexanone-d5	8.645	63	95590	105.77	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.77%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	118915	49.98	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.96%
66) 1,2-Dichlorobenzene-d4	12.201	152	89039	52.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.54%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
Data File : VU042278.D
Acq On : 05 Feb 2021 13:11
Operator : SY/MD
Sample : VU0205WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK110

Quant Time: Feb 06 07:24:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 06 07:13:30 2021
Response via : Initial Calibration

