

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
 Data File : VU042281.D
 Acq On : 05 Feb 2021 16:22
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: Feb 06 07:26:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 06 07:13:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	201958	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	184472	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	91559	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	71801	47.24	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 94.48%			
7) Chloroethane-d5	1.919	69	55259	53.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 106.04%			
11) 1,1-Dichloroethene-d2	2.575	63	121281	37.39	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 74.78%			
21) 2-Butanone-d5	4.642	46	162530	103.36	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 103.36%			
24) Chloroform-d	5.073	84	144717	47.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 94.82%			
26) 1,2-Dichloroethane-d4	5.713	65	105539	48.87	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 97.74%			
32) Benzene-d6	5.738	84	272703	50.88	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 101.76%			
36) 1,2-Dichloropropane-d6	6.700	67	91505	51.11	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 102.22%			
41) Toluene-d8	7.906	98	250611	49.36	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 98.72%			
43) trans-1,3-Dichloroprop...	8.189	79	46028	49.45	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 98.90%			
47) 2-Hexanone-d5	8.645	63	104992	103.72	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 103.72%			
56) 1,1,2,2-Tetrachloroeth...	10.767	84	126574	47.49	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 94.98%			
66) 1,2-Dichlorobenzene-d4	12.201	152	97273	50.67	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 101.34%			
Target Compounds						Qvalue
70) 1,2,4-trichlorobenzene	13.851	180	4778	2.57	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
Data File : VU042281.D
Acq On : 05 Feb 2021 16:22
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: Feb 06 07:26:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 06 07:13:30 2021
Response via : Initial Calibration

