

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
 Data File : VU042279.D
 Acq On : 05 Feb 2021 13:42
 Operator : SY/MD
 Sample : PB134469TB 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VLEB001

Quant Time: Feb 06 07:25:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 06 07:13:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	190201	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	176607	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.822	152	84825	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	67858	47.41	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	94.82%		
7) Chloroethane-d5	1.912	69	40802	41.56	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	83.12%		
11) 1,1-Dichloroethene-d2	2.568	63	112843	36.94	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.88%		
21) 2-Butanone-d5	4.645	46	147738	99.76	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	99.76%		
24) Chloroform-d	5.073	84	133624	46.48	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.96%		
26) 1,2-Dichloroethane-d4	5.713	65	99030	48.69	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.38%		
32) Benzene-d6	5.735	84	253555	49.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.82%		
36) 1,2-Dichloropropane-d6	6.700	67	85108	49.66	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.32%		
41) Toluene-d8	7.906	98	233904	48.12	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.24%		
43) trans-1,3-Dichloroprop...	8.189	79	42836	48.07	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	96.14%		
47) 2-Hexanone-d5	8.655	63	97450	100.55	ug/L	0.01
Spiked Amount 100.000	Range 45 - 130		Recovery =	100.55%		
56) 1,1,2,2-Tetrachloroeth...	10.770	84	122882	48.16	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	96.32%		
66) 1,2-Dichlorobenzene-d4	12.204	152	86631	48.71	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.42%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020521\
Data File : VU042279.D
Acq On : 05 Feb 2021 13:42
Operator : SY/MD
Sample : PB134469TB 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VLEB001

Quant Time: Feb 06 07:25:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020521WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 06 07:13:30 2021
Response via : Initial Calibration

