

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020623\
 Data File : VU052970.D
 Acq On : 06 Feb 2023 10:54
 Operator : JC/MD
 Sample : VU0206WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK175

Quant Time: Feb 06 13:42:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Feb 05 21:58:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	131949	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	129164	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	55080	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	55293	4.770	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.400%
7) Chloroethane-d5	1.912	69	51033	5.404	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.000%
11) 1,1-Dichloroethene-d2	2.565	65	24041	4.709	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.200%
20) 2-Butanone-d5	4.636	46	102151	32.957	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	65.920%
24) Chloroform-d	5.060	84	104574	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.700	65	51908	4.674	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	5.723	84	202522	4.838	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.687	67	68036	4.877	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.896	98	159380	4.318	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
43) trans-1,3-Dichloroprop...	8.176	79	19528	3.764	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.632	63	67839	31.478	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	62.960%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	46971	4.397	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	12.188	152	54550	5.440	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.044	84	5203	0.342	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020623\
Data File : VU052970.D
Acq On : 06 Feb 2023 10:54
Operator : JC/MD
Sample : VU0206WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK175

Quant Time: Feb 06 13:42:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Feb 05 21:58:05 2023
Response via : Initial Calibration

