

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020624\
 Data File : VU057816.D
 Acq On : 06 Feb 2024 11:54
 Operator : MD/SY
 Sample : P1174-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/16/2024
 Supervised By : Mahesh Dadoda 03/16/2024

Quant Time: Feb 12 12:25:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U020524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Feb 07 04:06:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.107	96	43071	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	14242	0.901	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	12291	0.879	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
Target Compounds						
8) 1,1,2-Trichloro-1,2,2-...	2.563	101	485453	45.920	ug/l	Qvalue 98
24) tert-Butyl Alcohol	3.239	59	9718m	16.744	ug/l	
25) Methyl tert-Butyl Ether	3.354	73	1439230	54.516	ug/l	96
31) Isopropyl Ether	3.988	45	1062632	30.617	ug/l	93
32) Ethyl-t-butyl ether	4.499	59	1418248	41.547	ug/l	97
33) Tert-Amyl methyl ether	5.940	73	1065152	37.272	ug/l	98
73) n-propylbenzene	10.901	120	592645	40.902	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020624\
Data File : VU057816.D
Acq On : 06 Feb 2024 11:54
Operator : MD/SY
Sample : P1174-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/16/2024
Supervised By : Mahesh Dadoda 03/16/2024

Quant Time: Feb 12 12:25:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U020524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Feb 07 04:06:29 2024
Response via : Initial Calibration

