

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020624\
 Data File : VU057822.D
 Acq On : 06 Feb 2024 15:40
 Operator : MD/SY
 Sample : P1174-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Feb 07 03:01:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U020524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 06 02:05:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.107	96	41066	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	13303	0.882	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
68) 1,2-Dichlorobenzene-d4	12.190	152	11878	0.891	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.033	84	6193	0.488	ug/l	92
27) Chloroform	5.078	83	357056	16.605	ug/l	98
44) Bromodichloromethane	7.103	83	46250	3.058	ug/l	98
55) Dibromochloromethane	8.804	129	104723	11.444	ug/l	99
67) Bromoform	10.286	173	74617	16.205	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020624\
Data File : VU057822.D
Acq On : 06 Feb 2024 15:40
Operator : MD/SY
Sample : P1174-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Feb 07 03:01:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U020524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 06 02:05:40 2024
Response via : Initial Calibration

