

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU020719\
 Data File : VU029364.D
 Acq On : 07 Feb 2019 09:54
 Operator : JC/SP
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 08 12:57:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 08 10:52:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.373	0.356	4.6	82	0.00
3 t	Chloromethane	0.341	0.309	9.4	80	0.00
4 Rt	Vinyl Chloride	0.400	0.383	4.3	80	0.00
5 T	Bromomethane	0.222	0.217	2.3	91	0.00
6 T	Chloroethane	0.234	0.230	1.7	84	0.00
7 T	Trichlorofluoromethane	0.558	0.591	-5.9	89	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.295	0.321	-8.8	91	0.00
9 Rt	1,1-Dichloroethene	0.288	0.306	-6.3	90	0.00
10 t	Iodomethane	0.370	0.485	-31.1#	95	0.00
11 t	Allyl Chloride	0.484	0.473	2.3	82	0.00
12 t	Acrylonitrile	0.065	0.064	1.5	84	0.00
13 T	Acetone	0.062	0.060	3.2	85	0.00
14 T	Carbon Disulfide	1.019	1.029	-1.0	86	0.00
15 RT	Methylene Chloride	0.350	0.332	5.1	88	0.00
16 RT	trans-1,2-Dichloroethene	0.323	0.336	-4.0	88	0.00
17 t	1,1-Dichloroethane	0.511	0.494	3.3	83	0.00
18 T	2-Butanone	0.070	0.068	2.9	76	0.00
19	Cyclohexane	0.474	0.427	9.9	81	0.00
20	Methylcyclohexane	0.498	0.505	-1.4	84	0.00
21 T	2,2-Dichloropropane	0.473	0.465	1.7	84	0.00
22 RT	cis-1,2-Dichloroethene	0.304	0.319	-4.9	87	0.00
23 t	Diethyl Ether	0.232	0.228	1.7	83	0.00
24 t	tert-Butyl Alcohol	0.024	0.023	4.2	83	0.00
25 t	Methyl tert-Butyl Ether	0.812	0.831	-2.3	86	0.00
26 t	Bromochloromethane	0.134	0.142	-6.0	89	0.00
27 t	Chloroform	0.497	0.510	-2.6	85	0.00
28 RT	1,1,1-Trichloroethane	0.461	0.476	-3.3	86	0.00
29 T	1,1-Dichloropropene	0.400	0.392	2.0	82	0.00
30 RT	Carbon Tetrachloride	0.417	0.443	-6.2	88	0.00
31 t	Isopropyl Ether	0.771	0.727	5.7	78	0.00
32	Ethyl-t-butyl ether	0.773	0.740	4.3	80	0.00
33	Tert-Amyl methyl ether	0.699	0.705	-0.9	83	0.00
34 t	Propionitrile	0.018	0.018	0.0	75	0.00
35 RT	Benzene	1.106	1.124	-1.6	84	0.00
36 RT	1,2-Dichloroethane	0.335	0.322	3.9	79	0.00
37 RT	Trichloroethene	0.317	0.347	-9.5	91	0.00
38 Rt	1,2-Dichloropropane	0.288	0.286	0.7	82	0.00
39 t	Methacrylonitrile	0.094	0.085	9.6	75	0.00
40 t	Methyl acrylate	0.142	0.136	4.2	78	0.00
41 t	Tetrahydrofuran	0.047	0.042	10.6	74	0.00
42 t	1-Chlorobutane	0.536	0.521	2.8	80	0.00
43 T	Dibromomethane	0.153	0.155	-1.3	86	0.00
44 T	Bromodichloromethane	0.391	0.398	-1.8	85	0.00
45 T	4-Methyl-2-Pentanone	0.180	0.171	5.0	76	0.00
46 t	t-1,4-Dichloro-2-butene	0.088	0.088	0.0	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU020719\
 Data File : VU029364.D
 Acq On : 07 Feb 2019 09:54
 Operator : JC/SP
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 08 12:57:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 08 10:52:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
47 t	Methyl methacrylate	0.135	0.138	-2.2	82	0.00
48 t	Ethyl methacrylate	0.289	0.292	-1.0	80	0.00
49 Rt	Toluene	0.708	0.736	-4.0	86	0.00
50 T	t-1,3-Dichloropropene	0.383	0.380	0.8	80	0.00
51 T	cis-1,3-Dichloropropene	0.455	0.456	-0.2	82	0.00
52 RT	1,1,2-Trichloroethane	0.203	0.210	-3.4	85	0.00
53 t	1,3-Dichloropropane	0.357	0.361	-1.1	83	0.00
54 t	2-Hexanone	0.126	0.121	4.0	75	0.00
55 t	Dibromochloromethane	0.289	0.306	-5.9	87	0.00
56 T	1,2-Dibromoethane	0.208	0.214	-2.9	86	0.00
57 S	4-Bromofluorobenzene	0.390	0.441	-13.1	96	0.00
58 RT	Tetrachloroethene	0.295	0.330	-11.9	92	0.00
59 Rt	Chlorobenzene	0.797	0.850	-6.6	88	0.00
60 T	1,1,1,2-Tetrachloroethane	0.295	0.318	-7.8	89	0.00
61 t	Pentachloroethane	0.228	0.250	-9.6	89	0.00
62 t	Hexachloroethane	0.249	0.273	-9.6	88	0.00
63 Rt	Ethyl Benzene	1.385	1.431	-3.3	85	0.00
64 RT	m/p-Xylenes	0.538	0.573	-6.5	86	0.00
65 RT	o-Xylene	0.528	0.564	-6.8	86	0.00
66 RT	Styrene	0.881	0.936	-6.2	85	0.00
67 t	Bromoform	0.180	0.199	-10.6	89	0.00
68 S	1,2-Dichlorobenzene-d4	0.400	0.444	-11.0	90	0.00
69 T	Isopropylbenzene	1.411	1.497	-6.1	86	0.00
70 T	1,1,2,2-Tetrachloroethane	0.260	0.265	-1.9	82	0.00
71 T	1,2,3-Trichloropropane	0.179	0.192	-7.3	83	0.00
72 t	Bromobenzene	0.342	0.385	-12.6	90	0.00
73 t	n-propylbenzene	0.393	0.434	-10.4	89	0.00
74 t	2-Chlorotoluene	0.335	0.369	-10.1	90	0.00
75 t	1,3,5-Trimethylbenzene	1.217	1.284	-5.5	86	0.00
76 t	4-Chlorotoluene	0.347	0.383	-10.4	89	0.00
77 t	tert-Butylbenzene	1.203	1.295	-7.6	87	0.00
78 t	1,2,4-Trimethylbenzene	1.232	1.309	-6.2	85	0.00
79 t	sec-Butylbenzene	1.574	1.667	-5.9	85	0.00
80	Nitrobenzene	0.012	0.012	0.0	67	0.00
81 t	p-Isopropyltoluene	1.334	1.450	-8.7	86	0.00
82 t	1,3-Dichlorobenzene	0.671	0.752	-12.1	90	0.00
83 Rt	1,4-Dichlorobenzene	0.668	0.744	-11.4	89	0.00
84 t	n-Butylbenzene	1.224	1.298	-6.0	83	0.00
85 Rt	1,2-Dichlorobenzene	0.635	0.693	-9.1	87	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.049	0.048	2.0	81	0.00
87 Rt	1,2,4-Trichlorobenzene	0.446	0.518	-16.1	89	0.00
88 t	Hexachlorobutadiene	0.258	0.312	-20.9	97	0.00
89 t	Naphthalene	0.756	0.849	-12.3	83	0.00
90 t	1,2,3-Trichlorobenzene	0.407	0.467	-14.7	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU020719\
Data File : VU029364.D
Acq On : 07 Feb 2019 09:54
Operator : JC/SP
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 08 12:57:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 08 10:52:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		