

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU020720\
 Data File : VU036695.D
 Acq On : 07 Feb 2020 12:04
 Operator : JC/MD
 Sample : VU0207MBL02
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK106

Quant Time: Feb 07 23:12:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM020620W.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 07 22:36:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	252109	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	236204	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.83	152	100798	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	74222	44.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.28%
7) Chloroethane-d5	1.93	69	71603	45.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.86%
11) 1,1-Dichloroethene-d2	2.59	63	107401	33.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.66%
21) 2-Butanone-d5	4.68	46	157969	110.13	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.13%
24) Chloroform-d	5.11	84	142276	46.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.52%
26) 1,2-Dichloroethane-d4	5.75	65	102176	49.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.18%
32) Benzene-d6	5.77	84	304480	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.06%
36) 1,2-Dichloropropane-d6	6.73	67	102599	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.54%
41) Toluene-d8	7.93	98	274516	48.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.54%
43) trans-1,3-Dichloropropene-	8.21	79	46708	47.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.64%
47) 2-Hexanone-d5	8.67	63	124151	105.52	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.52%
56) 1,1,2,2-Tetrachloroethane-	10.78	84	151084	49.73	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.46%
66) 1,2-Dichlorobenzene-d4	12.21	152	92317	50.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU020720\
Data File : VU036695.D
Acq On : 07 Feb 2020 12:04
Operator : JC/MD
Sample : VU0207MBL02
Misc : 5.00µ/5.0mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK106

Quant Time: Feb 07 23:12:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMULM020620W.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 07 22:36:50 2020
Response via : Initial Calibration

