

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020724\
 Data File : VU057835.D
 Acq On : 07 Feb 2024 13:09
 Operator : MD/SY
 Sample : VU0207MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Feb 08 01:26:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 08 01:18:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	250428	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	229661	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	108096	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	100673	50.225	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.460%		
7) Chloroethane-d5	1.894	69	76376	49.116	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	= 98.240%		
11) 1,1-Dichloroethene-d2	2.560	63	135110	36.701	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 73.400%		
21) 2-Butanone-d5	4.611	46	115324	103.670	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 103.670%		
24) Chloroform-d	5.052	84	164307	46.209	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 92.420%		
26) 1,2-Dichloroethane-d4	5.692	65	112141	49.751	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.500%		
32) Benzene-d6	5.717	84	346889	49.998	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.000%		
36) 1,2-Dichloropropane-d6	6.682	67	109810	50.118	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 100.240%		
41) Toluene-d8	7.891	98	310418	49.351	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 98.700%		
43) trans-1,3-Dichloroprop...	8.174	79	49727	46.395	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 92.780%		
47) 2-Hexanone-d5	8.624	63	91210	101.988	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 101.990%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	142962	50.124	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 100.240%		
66) 1,2-Dichlorobenzene-d4	12.187	152	107330	51.682	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 103.360%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020724\
Data File : VU057835.D
Acq On : 07 Feb 2024 13:09
Operator : MD/SY
Sample : VU0207MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Feb 08 01:26:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020724WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 08 01:18:21 2024
Response via : Initial Calibration

