

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020724\
 Data File : VU057845.D
 Acq On : 07 Feb 2024 17:11
 Operator : MD/SY
 Sample : P1329-19ME
 Misc : 4.51g/5.0ml/100uL//5.0mL/MSVOA_U/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FM4ME

Quant Time: Feb 08 01:29:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 08 01:18:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	256351	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	250225	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	127213	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	90165	43.943	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.880%
7) Chloroethane-d5	1.869	69	23882	15.003	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	30.000%#
11) 1,1-Dichloroethene-d2	2.534	63	117692	31.231	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.460%
21) 2-Butanone-d5	4.621	46	88160	77.420	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.420%
24) Chloroform-d	5.049	84	148976	40.930	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.860%
26) 1,2-Dichloroethane-d4	5.689	65	100829	43.699	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.400%
32) Benzene-d6	5.711	84	313744	41.504	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.679	67	99167	41.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.080%
41) Toluene-d8	7.891	98	278508	40.639	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.280%
43) trans-1,3-Dichloroprop...	8.177	79	48288	41.350	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.700%
47) 2-Hexanone-d5	8.643	63	97361	99.919	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.920%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	150165	48.323	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.640%
66) 1,2-Dichlorobenzene-d4	12.190	152	112276	45.939	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020724\
Data File : VU057845.D
Acq On : 07 Feb 2024 17:11
Operator : MD/SY
Sample : P1329-19ME
Misc : 4.51g/5.0ml/100uL//5.0mL/MSVOA_U/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FM4ME

Quant Time: Feb 08 01:29:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM020724WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 08 01:18:21 2024
Response via : Initial Calibration

