

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU020819\
 Data File : VU029375.D
 Acq On : 08 Feb 2019 12:30
 Operator : JC/SP
 Sample : K1161-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Feb 18 12:07:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 08 10:52:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	55175	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	20638	0.96	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	21641	0.98	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.00%	
Target Compounds						Ovalue
8) 1,1,2-Trichloro-1,2,2-trif	2.29	101	835674	51.379	ug/l	89
24) tert-Butyl Alcohol	2.84	59	58503	44.044	ug/l #	71
25) Methyl tert-Butyl Ether	3.00	73	1844841	41.164	ug/l #	86
31) Isopropyl Ether	3.58	45	1842052	43.280	ug/l	91
32) Ethyl-t-butyl ether	4.07	59	1765936	41.405	ug/l	88
33) Tert-Amyl methyl ether	5.58	73	2002510	51.925	ug/l	98
73) n-propylbenzene	10.58	120	925669	42.679	ug/l	78
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU020819\
Data File : VU029375.D
Acq On : 08 Feb 2019 12:30
Operator : JC/SP
Sample : K1161-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Feb 18 12:07:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 08 10:52:46 2019
Response via : Initial Calibration

