

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU020819\
 Data File : VU029376.D
 Acq On : 08 Feb 2019 13:34
 Operator : JC/SP
 Sample : K1161-10DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WSDL

Quant Time: Feb 09 03:35:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 08 10:52:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	58984	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	21690	0.94	ug/l	0.00
Spiked Amount 1.000			Recovery	=	94.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	23613	1.00	ug/l	0.00
Spiked Amount 1.000			Recovery	=	100.00%	
Target Compounds						
					Ovalue	
8) 1,1,2-Trichloro-1,2,2-trif	2.29	101	170824	9.824	ug/l	88
24) tert-Butyl Alcohol	2.84	59	13733	9.671	ug/l #	71
25) Methyl tert-Butyl Ether	3.00	73	374554	7.818	ug/l #	86
31) Isopropyl Ether	3.58	45	377584	8.299	ug/l	91
32) Ethyl-t-butyl ether	4.07	59	354238	7.769	ug/l #	87
33) Tert-Amyl methyl ether	5.58	73	403416	9.785	ug/l	98
73) n-propylbenzene	10.58	120	182233	7.859	ug/l	79
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU020819\
Data File : VU029376.D
Acq On : 08 Feb 2019 13:34
Operator : JC/SP
Sample : K1161-10DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Quant Time: Feb 09 03:35:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 08 10:52:46 2019
Response via : Initial Calibration

