

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU020819\
 Data File : VU029378.D
 Acq On : 08 Feb 2019 14:34
 Operator : JC/SP
 Sample : K1161-09DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Feb 09 03:41:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Feb 08 10:52:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	55788	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	19925	0.91	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	20144	0.90	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.00%	
Target Compounds						Ovalue
27) Chloroform	4.67	83	107033	3.859	ug/l	97
44) Bromodichloromethane	6.76	83	201805	9.257	ug/l	95
55) Dibromochloromethane	8.48	129	20096	1.248	ug/l	98
67) Bromoform	9.96	173	54273	5.419	ug/l	100
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU020819\
Data File : VU029378.D
Acq On : 08 Feb 2019 14:34
Operator : JC/SP
Sample : K1161-09DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Feb 09 03:41:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U020519DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Feb 08 10:52:46 2019
Response via : Initial Calibration

