

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020921\
 Data File : VU042299.D
 Acq On : 09 Feb 2021 16:04
 Operator : SY/MD
 Sample : M1351-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR7

Quant Time: Feb 10 02:31:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Feb 10 02:29:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	99374	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	94196	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	48358	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	28462	4.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.80%
7) Chloroethane-d5	1.919	69	25073	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.575	63	52506	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.645	46	132921	56.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.66%
24) Chloroform-d	5.076	84	67780	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.713	65	43103	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
32) Benzene-d6	5.739	84	125655	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.700	67	42075	5.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.00%
41) Toluene-d8	7.906	98	112529	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloroprop...	8.189	79	17221	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.645	63	91785	52.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.84%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	35012	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	12.201	152	45279	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020921\
Data File : VU042299.D
Acq On : 09 Feb 2021 16:04
Operator : SY/MD
Sample : M1351-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR7

Quant Time: Feb 10 02:31:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Feb 10 02:29:25 2021
Response via : Initial Calibration

