

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
 Data File : VU053015.D
 Acq On : 09 Feb 2023 13:26
 Operator : JC/MD
 Sample : VU0209WBS01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01

Quant Time: Feb 09 13:46:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.112	96	33216	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.630	95	12904	1.168	ug/l	0.00
Spiked Amount 1.000			Recovery	=	117.000%	
68) 1,2-Dichlorobenzene-d4	12.189	152	10824	0.952	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.000%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.383	85	19290	1.716	ug/l	96
3) Chloromethane	1.518	50	19982	1.873	ug/l	99
4) Vinyl Chloride	1.605	62	21968	2.133	ug/l	98
5) Bromomethane	1.859	94	12833	2.397	ug/l	94
6) Chloroethane	1.936	64	13064	2.109	ug/l	97
7) Trichlorofluoromethane	2.138	101	33493	2.232	ug/l	99
8) 1,1,2-Trichloro-1,2,2-...	2.579	101	19142	2.333	ug/l	92
9) 1,1-Dichloroethene	2.579	96	17353	2.239	ug/l	94
10) Iodomethane	2.720	142	20935	2.440	ug/l	96
11) Allyl Chloride	2.920	41	23447	2.532	ug/l	99
12) Acrylonitrile	3.325	53	8102	4.930	ug/l #	91
13) Acetone	2.659	43	35679	15.308	ug/l	92
14) Carbon Disulfide	2.791	76	50835	1.958	ug/l	98
15) Methylene Chloride	3.042	84	20344	2.173	ug/l	96
16) trans-1,2-Dichloroethene	3.347	96	19147	2.220	ug/l	96
17) 1,1-Dichloroethane	3.865	63	37473	2.574	ug/l	95
18) 2-Butanone	4.733	43	38932	14.378	ug/l	89
19) Cyclohexane	5.379	56	28950	2.261	ug/l	96
20) Methylcyclohexane	6.759	83	30195	2.113	ug/l	98
21) 2,2-Dichloropropane	4.659	77	33980	2.513	ug/l	98
22) cis-1,2-Dichloroethene	4.665	96	20902	2.319	ug/l	92
23) Diethyl Ether	2.376	59	13468	2.458	ug/l	96
24) tert-Butyl Alcohol	3.318	59	2590	3.926	ug/l #	1
25) Methyl tert-Butyl Ether	3.363	73	49255	2.770	ug/l	96
26) Bromochloromethane	4.971	128	8924	2.121	ug/l	84
27) Chloroform	5.087	83	37041	2.373	ug/l	100
28) 1,1,1-Trichloroethane	5.312	97	34229	2.375	ug/l	98
29) 1,1-Dichloropropene	5.521	75	24826	2.043	ug/l	99
30) Carbon Tetrachloride	5.518	117	23618	1.806	ug/l	97
31) Isopropyl Ether	3.994	45	54025	2.581	ug/l	99
32) Ethyl-t-butyl ether	4.508	59	55303	2.830	ug/l	96
33) Tert-Amyl methyl ether	5.945	73	43075	2.492	ug/l	98
34) Propionitrile	4.820	54	6846	11.042	ug/l #	2
35) Benzene	5.768	78	67875	1.925	ug/l	98
36) 1,2-Dichloroethane	5.794	62	19751	1.865	ug/l	95
37) Trichloroethene	6.540	130	17383	1.808	ug/l	90
38) 1,2-Dichloropropane	6.791	63	18942	2.117	ug/l	96
39) Methacrylonitrile	4.984	41	5889	2.378	ug/l #	88
40) Methyl acrylate	4.852	55	11134	2.537	ug/l	98
41) Tetrahydrofuran	5.074	42	6924	5.621	ug/l #	71
42) 1-Chlorobutane	5.453	56	32309	2.080	ug/l	99
43) Dibromomethane	6.919	93	8990	1.985	ug/l	85
44) Bromodichloromethane	7.103	83	24838	2.022	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
 Data File : VU053015.D
 Acq On : 09 Feb 2023 13:26
 Operator : JC/MD
 Sample : VU0209WBS01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01

Quant Time: Feb 09 13:46:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.797	43	52295	10.990	ug/l	95
46) t-1,4-Dichloro-2-butene	10.832	75	22019	8.826	ug/l #	57
47) Methyl methacrylate	6.964	69	17865	4.438	ug/l	97
48) Ethyl methacrylate	8.334	69	18152	2.488	ug/l	97
49) Toluene	7.968	92	42875	2.003	ug/l	94
50) t-1,3-Dichloropropene	8.209	75	24086	2.248	ug/l	96
51) cis-1,3-Dichloropropene	7.607	75	28363	2.250	ug/l	96
52) 1,1,2-Trichloroethane	8.398	97	13410	2.083	ug/l	96
53) 1,3-Dichloropropane	8.575	76	22591	2.030	ug/l	98
54) 2-Hexanone	8.688	43	50892	12.225	ug/l	96
55) Dibromochloromethane	8.807	129	15376	1.826	ug/l	97
56) 1,2-Dibromoethane	8.923	107	12147	1.981	ug/l	97
58) Tetrachloroethene	8.550	164	13790	1.590	ug/l	91
59) Chlorobenzene	9.443	112	45681	1.961	ug/l	98
60) 1,1,1,2-Tetrachloroethane	9.533	131	16025	1.777	ug/l	94
61) Pentachloroethane	11.424	117	13605	1.841	ug/l	88
62) Hexachloroethane	12.469	117	14800	1.983	ug/l #	77
63) Ethyl Benzene	9.566	91	85731	2.271	ug/l	99
64) m/p-Xylenes	9.691	106	62246	4.137	ug/l	92
65) o-Xylene	10.096	106	31820	2.191	ug/l	95
66) Styrene	10.112	104	49072	2.100	ug/l	94
67) Bromoform	10.289	173	8111	1.632	ug/l #	97
69) Isopropylbenzene	10.482	105	82247	2.226	ug/l	99
70) 1,1,2,2-Tetrachloroethane	10.781	83	17575	2.190	ug/l #	96
71) 1,2,3-Trichloropropane	10.832	75	22019	3.169	ug/l #	59
72) Bromobenzene	10.781	156	16261	1.733	ug/l	83
73) n-propylbenzene	10.903	120	21676	2.103	ug/l	90
74) 2-Chlorotoluene	10.984	126	18313	1.962	ug/l	87
75) 1,3,5-Trimethylbenzene	11.083	105	69077	2.164	ug/l	94
76) 4-Chlorotoluene	11.093	126	19503	2.036	ug/l	92
77) tert-Butylbenzene	11.418	119	66661	2.075	ug/l	92
78) 1,2,4-Trimethylbenzene	11.463	105	70590	2.190	ug/l	94
79) sec-Butylbenzene	11.640	105	90796	2.173	ug/l	95
80) Nitrobenzene	13.209	77	5619	11.309	ug/l #	91
81) p-Isopropyltoluene	11.787	119	72231	2.074	ug/l #	94
82) 1,3-Dichlorobenzene	11.742	146	34483	1.800	ug/l	95
83) 1,4-Dichlorobenzene	11.832	146	34219	1.779	ug/l #	93
84) n-Butylbenzene	12.202	91	71416	2.329	ug/l	97
85) 1,2-Dichlorobenzene	12.209	146	31571	1.746	ug/l	93
86) 1,2-Dibromo-3-Chloropr...	12.990	75	3136	2.109	ug/l #	80
87) 1,2,4-Trichlorobenzene	13.836	180	20672	1.906	ug/l	95
88) Hexachlorobutadiene	14.016	225	10953	1.441	ug/l	99
89) Naphthalene	14.083	128	45708	2.174	ug/l	99
90) 1,2,3-Trichlorobenzene	14.324	180	19249	1.870	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\  
Data File : VU053015.D  
Acq On    : 09 Feb 2023  13:26  
Operator  : JC/MD  
Sample    : VU0209WBS01  
Misc      : 25.0mL/MSVOA_U/WATER  
ALS Vial  : 5    Sample Multiplier: 1
```

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Quant Time: Feb 09 13:46:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.112 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 96	Resp: 33216
Ion Ratio	Lower Upper
96 100	
70 21.1	16.2 24.4
95 9.0	7.4 11.0
97 7.3	0.0 0.0#



#2
Dichlorodifluoromethane
Concen: 1.716 ug/l
RT: 1.383 min Scan# 91
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 85	Resp: 19290
Ion Ratio	Lower Upper
85 100	
87 29.7	16.1 48.2









#3
Chloromethane
Concen: 1.873 ug/l
RT: 1.518 min Scan# 11
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 50 Resp: 19982
Ion Ratio Lower Upper
50 100
52 34.0 26.8 40.2



#4
Vinyl Chloride
Concen: 2.133 ug/l
RT: 1.605 min Scan# 160
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 62 Resp: 21968
Ion Ratio Lower Upper
62 100
64 31.6 26.0 39.0









#5
Bromomethane
Concen: 2.397 ug/l
RT: 1.859 min Scan# 21
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 94 Resp: 12833
Ion Ratio Lower Upper
94 100
96 98.4 73.8 110.6



#6
Chloroethane
Concen: 2.109 ug/l
RT: 1.936 min Scan# 263
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 64 Resp: 13064
Ion Ratio Lower Upper
64 100
66 35.3 26.6 40.0









#7
 Trichlorofluoromethane
 Concen: 2.232 ug/l
 RT: 2.138 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26




Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01


Tgt Ion:101 Resp: 33493
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.6 77.4



#8
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 2.333 ug/l
 RT: 2.579 min Scan# 463
 Delta R.T. -0.003 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26

Tgt Ion:101 Resp: 19142
 Ion Ratio Lower Upper
 101 100
 85 45.4 34.1 51.1
 151 69.6 63.0 94.4









#9
1,1-Dichloroethene
Concen: 2.239 ug/l
RT: 2.579 min Scan# 41
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 96 Resp: 17353
Ion Ratio Lower Upper
96 100
61 161.1 0.0 460.5
98 59.5 0.0 128.0



#10
Iodomethane
Concen: 2.440 ug/l
RT: 2.720 min Scan# 507
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 142 Resp: 20935
Ion Ratio Lower Upper
142 100
127 46.7 35.4 53.0









#11
 Allyl Chloride
 Concen: 2.532 ug/l
 RT: 2.920 min Scan# 51
 Delta R.T. -0.003 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26




Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01


Tgt Ion: 41 Resp: 23447
 Ion Ratio Lower Upper
 41 100
 39 73.1 59.3 88.9



#12
 Acrylonitrile
 Concen: 4.930 ug/l
 RT: 3.325 min Scan# 695
 Delta R.T. 0.004 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26

Tgt Ion: 53 Resp: 8102
 Ion Ratio Lower Upper
 53 100
 52 84.8 63.0 94.4
 51 30.3 31.3 46.9#





#13

Acetone

Concen: 15.308 ug/l

RT: 2.659 min Scan# 41

Delta R.T. 0.007 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :





VU0209WBS01

Tgt Ion: 43 Resp: 35679

Ion Ratio Lower Upper

43 100

58 40.6 29.0 43.4



#14

Carbon Disulfide

Concen: 1.958 ug/l

RT: 2.791 min Scan# 529

Delta R.T. -0.003 min

Lab File: VU053015.D




Acq: 09 Feb 2023 13:26


Tgt Ion: 76 Resp: 50835

Ion Ratio Lower Upper

76 100

78 9.7 7.1 10.7









#15
Methylene Chloride
Concen: 2.173 ug/l
RT: 3.042 min Scan# 60
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 84 Resp: 20344
Ion Ratio Lower Upper
84 100
49 114.2 86.7 130.1
51 36.3 0.0 66.8
86 63.7 51.5 77.3



#16
trans-1,2-Dichloroethene
Concen: 2.220 ug/l
RT: 3.347 min Scan# 702
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 96 Resp: 19147
Ion Ratio Lower Upper
96 100
61 136.4 105.7 158.5
98 59.7 51.4 77.2









#17
1,1-Dichloroethane
Concen: 2.574 ug/l
RT: 3.865 min Scan# 80
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 63 Resp: 37473
Ion Ratio Lower Upper
63 100
98 5.3 3.8 11.2
100 3.7 2.2 6.6



#18
2-Butanone
Concen: 14.378 ug/l
RT: 4.733 min Scan# 1133
Delta R.T. 0.007 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 43 Resp: 38932
Ion Ratio Lower Upper
43 100
57 12.8 0.0 17.4









#19
Cyclohexane
Concen: 2.261 ug/l
RT: 5.379 min Scan# 11763
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Tgt Ion: 56 Resp: 28950




Ion	Ratio	Lower	Upper
56	100		
69	35.6	26.8	40.2
84	92.8	77.0	115.6




#20
Methylcyclohexane
Concen: 2.113 ug/l
RT: 6.759 min Scan# 1763
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 83 Resp: 30195

Ion	Ratio	Lower	Upper
83	100		
55	72.1	58.7	88.1
98	47.6	39.4	59.2









#21
2,2-Dichloropropane
Concen: 2.513 ug/l
RT: 4.659 min Scan# 1110
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 77 Resp: 33980
Ion Ratio Lower Upper
77 100
97 21.7 16.8 25.2



#22
cis-1,2-Dichloroethene
Concen: 2.319 ug/l
RT: 4.665 min Scan# 1112
Delta R.T. 0.004 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 96 Resp: 20902
Ion Ratio Lower Upper
96 100
61 153.9 0.0 353.3
98 66.7 32.6 97.7









#23
Diethyl Ether
Concen: 2.458 ug/l
RT: 2.376 min Scan# 40
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 59 Resp: 13468
Ion Ratio Lower Upper
59 100
45 65.7 52.1 78.1
74 82.7 61.6 92.4



#24
tert-Butyl Alcohol
Concen: 3.926 ug/l
RT: 3.318 min Scan# 693
Delta R.T. 0.030 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 59 Resp: 2590
Ion Ratio Lower Upper
59 100
57 404.4 45.8 68.8#









#25
Methyl tert-Butyl Ether
Concen: 2.770 ug/l
RT: 3.363 min Scan# 70
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 73 Resp: 49255
Ion Ratio Lower Upper
73 100
43 17.4 15.3 22.9



#26
Bromochloromethane
Concen: 2.121 ug/l
RT: 4.971 min Scan# 1207
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:128 Resp: 8924
Ion Ratio Lower Upper
128 100
49 162.2 0.0 265.0
130 119.1 100.7 151.1





#27

Chloroform

Concen: 2.373 ug/l

RT: 5.087 min Scan# 1165

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :





VU0209WBS01

Tgt Ion: 83 Resp: 37041

Ion Ratio Lower Upper

83 100

85 65.0 0.0 130.6



#28

1,1,1-Trichloroethane

Concen: 2.375 ug/l

RT: 5.312 min Scan# 1313

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26




Tgt Ion: 97 Resp: 34229


Ion Ratio Lower Upper

97 100

99 63.5 31.9 95.5

61 43.8 20.8 62.2









#29
1,1-Dichloropropene
Concen: 2.043 ug/l
RT: 5.521 min Scan# 11
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Tgt Ion: 75 Resp: 24826




Ion	Ratio	Lower	Upper
75	100		
110	33.9	17.4	52.2
77	30.6	24.8	37.2




#30
Carbon Tetrachloride
Concen: 1.806 ug/l
RT: 5.518 min Scan# 1377
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 117 Resp: 23618

Ion	Ratio	Lower	Upper
117	100		
119	99.4	76.4	114.6
121	30.3	25.0	37.6









#31
Isopropyl Ether
Concen: 2.581 ug/l
RT: 3.994 min Scan# 90
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 45 Resp: 54025
Ion Ratio Lower Upper
45 100
43 49.4 24.5 73.5



#32
Ethyl-t-butyl ether
Concen: 2.830 ug/l
RT: 4.508 min Scan# 1063
Delta R.T. 0.004 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 59 Resp: 55303
Ion Ratio Lower Upper
59 100
87 42.5 36.2 54.2









#33
Tert-Amyl methyl ether
Concen: 2.492 ug/l
RT: 5.945 min Scan# 1160
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 73	Resp: 43075
Ion Ratio	Lower Upper
73 100	
87 23.7	20.1 30.1
71 11.4	9.2 13.8



#34
Propionitrile
Concen: 11.042 ug/l
RT: 4.820 min Scan# 1160
Delta R.T. 0.014 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt	Ion: 54	Resp:	6846
Ion	Ratio	Lower	Upper
54	100		
52	16.8	16.9	25.3#
55	5.6	118.6	178.0#
40	1.3	2.8	4.2#









#35
Benzene
Concen: 1.925 ug/l
RT: 5.768 min Scan# 14
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 78 Resp: 67875
Ion Ratio Lower Upper
78 100
77 22.9 19.0 28.4



#36
1,2-Dichloroethane
Concen: 1.865 ug/l
RT: 5.794 min Scan# 1463
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 62 Resp: 19751
Ion Ratio Lower Upper
62 100
98 11.2 7.5 11.3









#37
Trichloroethene
Concen: 1.808 ug/l
RT: 6.540 min Scan# 1773
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:130 Resp: 17383
Ion Ratio Lower Upper
130 100
95 105.6 76.6 114.8



#38
1,2-Dichloropropane
Concen: 2.117 ug/l
RT: 6.791 min Scan# 1773
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 63 Resp: 18942
Ion Ratio Lower Upper
63 100
65 32.8 24.6 36.8





#39

Methacrylonitrile

Concen: 2.378 ug/l

RT: 4.984 min Scan# 1131

Delta R.T. 0.007 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :

VU0209WBS01

Tgt Ion: 41 Resp: 5889





Ion Ratio Lower Upper

41 100

67 92.3 64.9 97.3

39 54.0 41.7 62.5

52 45.5 26.0 39.0#



#40

Methyl acrylate

Concen: 2.537 ug/l

RT: 4.852 min Scan# 1170

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Tgt Ion: 55 Resp: 11134




Ion Ratio Lower Upper


55 100

85 17.5 14.4 21.6

58 8.2 7.3 10.9

42 10.0 7.6 11.4





#41

Tetrahydrofuran

Concen: 5.621 ug/l

RT: 5.074 min Scan# 1161

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :

VU0209WBS01





Tgt Ion: 42 Resp: 6924

Ion Ratio Lower Upper

42 100

72 52.6 49.1 73.7

71 19.6 43.8 65.6#



#42

1-Chlorobutane

Concen: 2.080 ug/l

RT: 5.453 min Scan# 1357

Delta R.T. -0.003 min

Lab File: VU053015.D




Acq: 09 Feb 2023 13:26


Tgt Ion: 56 Resp: 32309

Ion Ratio Lower Upper

56 100

41 48.1 24.5 73.5









#43
Dibromomethane
Concen: 1.985 ug/l
RT: 6.919 min Scan# 11
Delta R.T. 0.004 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Tgt Ion: 93 Resp: 8990




Ion	Ratio	Lower	Upper
93	100		
95	91.7	67.0	100.6
174	82.3	82.2	123.4




#44
Bromodichloromethane
Concen: 2.022 ug/l
RT: 7.103 min Scan# 1870
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 83 Resp: 24838

Ion	Ratio	Lower	Upper
83	100		
85	66.8	52.1	78.1
127	8.2	7.2	10.8









#45
4-Methyl-2-Pentanone
Concen: 10.990 ug/l
RT: 7.797 min Scan# 2008
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:	43	Resp:	52295
Ion	Ratio	Lower	Upper
43	100		
58	47.0	21.7	65.1
85	21.6	16.2	24.4



#46
t-1,4-Dichloro-2-butene
Concen: 8.826 ug/l
RT: 10.832 min Scan# 3030
Delta R.T. 0.004 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:	75	Resp:	22019
Ion	Ratio	Lower	Upper
75	100		
53	31.4	57.0	85.4#
89	23.2	41.1	61.7#
88	16.6	32.3	48.5#









#47
Methyl methacrylate
Concen: 4.438 ug/l
RT: 6.964 min Scan# 11
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Tgt Ion: 69 Resp: 17865




Ion	Ratio	Lower	Upper
69	100		
41	100.9	0.0	201.8
39	58.5	44.0	66.0
100	34.0	31.5	47.3




#48
Ethyl methacrylate
Concen: 2.488 ug/l
RT: 8.334 min Scan# 2253
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 69 Resp: 18152

Ion	Ratio	Lower	Upper
69	100		
41	58.7	28.1	84.4





#49

Toluene

Concen: 2.003 ug/l

RT: 7.968 min Scan# 2119

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :





VU0209WBS01

Tgt Ion: 92 Resp: 42875

Ion Ratio Lower Upper

92 100

91 176.0 133.8 200.8



#50

t-1,3-Dichloropropene

Concen: 2.248 ug/l

RT: 8.209 min Scan# 2214

Delta R.T. 0.001 min

Lab File: VU053015.D




Acq: 09 Feb 2023 13:26


Tgt Ion: 75 Resp: 24086

Ion Ratio Lower Upper

75 100

77 32.4 24.4 36.6









#51
 cis-1,3-Dichloropropene
 Concen: 2.250 ug/l
 RT: 7.607 min Scan# 20
 Delta R.T. 0.001 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26




Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01


Tgt Ion:	75	Resp:	28363
Ion	Ratio	Lower	Upper
75	100		
77	28.5	25.5	38.3
39	39.0	32.3	48.5



#52
 1,1,2-Trichloroethane
 Concen: 2.083 ug/l
 RT: 8.398 min Scan# 2273
 Delta R.T. 0.001 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26

Tgt Ion:	97	Resp:	13410
Ion	Ratio	Lower	Upper
97	100		
83	87.0	68.2	102.2
85	50.6	44.5	66.7
99	60.2	51.0	76.4





#53

1,3-Dichloropropane

Concen: 2.030 ug/l

RT: 8.575 min Scan# 21

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :





VU0209WBS01

Tgt Ion: 76 Resp: 22591

Ion Ratio Lower Upper

76 100

78 32.9 25.2 37.8



#54

2-Hexanone

Concen: 12.225 ug/l

RT: 8.688 min Scan# 2363

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26




Tgt Ion: 43 Resp: 50892


Ion Ratio Lower Upper

43 100

58 63.5 40.0 80.0

57 22.0 0.6 40.6









#55
Dibromochloromethane
Concen: 1.826 ug/l
RT: 8.807 min Scan# 2436
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:129 Resp: 15376
Ion Ratio Lower Upper
129 100
127 79.4 61.5 92.3



#56
1,2-Dibromoethane
Concen: 1.981 ug/l
RT: 8.923 min Scan# 2436
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:107 Resp: 12147
Ion Ratio Lower Upper
107 100
109 92.7 0.0 190.4





#57

4-Bromofluorobenzene

Concen: 1.168 ug/l

RT: 10.630 min Scan# 2967

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :

VU0209WBS01





Tgt Ion: 95 Resp: 12904

Ion Ratio Lower Upper

95 100

174 62.6 66.9 100.3#

176 63.2 61.9 92.9



#58

Tetrachloroethene

Concen: 1.590 ug/l

RT: 8.550 min Scan# 2320

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Tgt Ion: 164 Resp: 13790




Ion Ratio Lower Upper


164 100

166 129.7 101.5 152.3

129 108.7 75.2 112.8

131 101.1 72.2 108.2





#59

Chlorobenzene

Concen: 1.961 ug/l

RT: 9.443 min Scan# 21

Delta R.T. 0.001 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26

Instrument :

MSVOA_U

ClientSampleId :





VU0209WBS01

Tgt Ion:112 Resp: 45681

Ion Ratio Lower Upper

112 100

114 32.6 25.2 37.8



#60

1,1,1,2-Tetrachloroethane

Concen: 1.777 ug/l

RT: 9.533 min Scan# 2626

Delta R.T. 0.004 min

Lab File: VU053015.D

Acq: 09 Feb 2023 13:26




Tgt Ion:131 Resp: 16025


Ion Ratio Lower Upper

131 100

133 92.1 77.2 115.8

119 69.0 50.8 76.2









#61
Pentachloroethane
Concen: 1.841 ug/l
RT: 11.424 min Scan# 3136
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:117 Resp: 13605
Ion Ratio Lower Upper
117 100
167 73.0 66.8 100.2



#62
Hexachloroethane
Concen: 1.983 ug/l
RT: 12.469 min Scan# 3539
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:117 Resp: 14800
Ion Ratio Lower Upper
117 100
201 60.4 64.6 97.0#









#63
Ethyl Benzene
Concen: 2.271 ug/l
RT: 9.566 min Scan# 2636
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 91 Resp: 85731
Ion Ratio Lower Upper
91 100
106 30.7 25.1 37.7



#64
m/p-Xylenes
Concen: 4.137 ug/l
RT: 9.691 min Scan# 2675
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 106 Resp: 62246
Ion Ratio Lower Upper
106 100
91 209.9 158.7 238.1









#65
o-Xylene
Concen: 2.191 ug/l
RT: 10.096 min Scan# 21
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:106 Resp: 31820
Ion Ratio Lower Upper
106 100
91 218.7 105.6 316.8



#66
Styrene
Concen: 2.100 ug/l
RT: 10.112 min Scan# 2806
Delta R.T. 0.000 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:104 Resp: 49072
Ion Ratio Lower Upper
104 100
78 54.6 39.6 59.4
103 56.8 42.9 64.3









#67
Bromoform
Concen: 1.632 ug/l
RT: 10.289 min Scan# 21
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:173 Resp: 8111
Ion Ratio Lower Upper
173 100
175 47.3 39.5 59.3
254 7.9 8.0 12.0#



#68
1,2-Dichlorobenzene-d4
Concen: 0.952 ug/l
RT: 12.189 min Scan# 3452
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:152 Resp: 10824
Ion Ratio Lower Upper
152 100
115 80.6 0.0 262.6
150 185.6 0.0 638.0









#69
Isopropylbenzene
Concen: 2.226 ug/l
RT: 10.482 min Scan# 2921
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:105 Resp: 82247
Ion Ratio Lower Upper
105 100
120 26.3 13.3 39.9



#70
1,1,2,2-Tetrachloroethane
Concen: 2.190 ug/l
RT: 10.781 min Scan# 3014
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 83 Resp: 17575
Ion Ratio Lower Upper
83 100
131 8.6 8.6 13.0#
85 62.9 52.6 78.8









#71
1,2,3-Trichloropropane
Concen: 3.169 ug/l
RT: 10.832 min Scan# 3014
Delta R.T. 0.010 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion: 75	Resp: 22019
Ion Ratio	Lower Upper
75 100	
77 0.0	0.0 86.0
110 19.6	0.0 63.8
97 11.3	0.0 31.8



#72
Bromobenzene
Concen: 1.733 ug/l
RT: 10.781 min Scan# 3014
Delta R.T. 0.004 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt	Ion:156	Resp:	16261
Ion	Ratio	Lower	Upper
156	100		
77	194.4	0.0	321.2
158	99.0	0.0	192.0









#73
n-propylbenzene
Concen: 2.103 ug/l
RT: 10.903 min Scan# 3077
Delta R.T. 0.000 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:120 Resp: 21676
Ion Ratio Lower Upper
120 100
91 460.8 348.0 522.0



#74
2-Chlorotoluene
Concen: 1.962 ug/l
RT: 10.984 min Scan# 3077
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:126 Resp: 18313
Ion Ratio Lower Upper
126 100
91 319.9 0.0 588.6









#75
1,3,5-Trimethylbenzene
Concen: 2.164 ug/l
RT: 11.083 min Scan# 3110
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:105 Resp: 69077
Ion Ratio Lower Upper
105 100
120 45.8 39.9 59.9



#76
4-Chlorotoluene
Concen: 2.036 ug/l
RT: 11.093 min Scan# 3111
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:126 Resp: 19503
Ion Ratio Lower Upper
126 100
91 346.8 0.0 661.8









#77
tert-Butylbenzene
Concen: 2.075 ug/l
RT: 11.418 min Scan# 3134
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:119 Resp: 66661
Ion Ratio Lower Upper
119 100
91 64.3 28.4 85.3
134 21.7 18.6 27.8



#78
1,2,4-Trimethylbenzene
Concen: 2.190 ug/l
RT: 11.463 min Scan# 3226
Delta R.T. -0.002 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:105 Resp: 70590
Ion Ratio Lower Upper
105 100
120 43.2 23.4 70.3









#79
 sec-Butylbenzene
 Concen: 2.173 ug/l
 RT: 11.640 min Scan# 31
 Delta R.T. 0.001 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26




Instrument :
 MSVOA_U
 ClientSampleId :
 VU0209WBS01


Tgt Ion:105 Resp: 90796
 Ion Ratio Lower Upper
 105 100
 134 18.5 16.5 24.7



#80
 Nitrobenzene
 Concen: 11.309 ug/l
 RT: 13.209 min Scan# 3769
 Delta R.T. 0.008 min
 Lab File: VU053015.D
 Acq: 09 Feb 2023 13:26

Tgt Ion: 77 Resp: 5619
 Ion Ratio Lower Upper
 77 100
 123 36.6 18.8 66.8
 65 13.2 10.1 12.9#









#81
p-Isopropyltoluene
Concen: 2.074 ug/l
RT: 11.787 min Scan# 311
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:119 Resp: 72231
Ion Ratio Lower Upper
119 100
134 24.4 20.7 31.1
91 26.7 17.6 26.4#



#82
1,3-Dichlorobenzene
Concen: 1.800 ug/l
RT: 11.742 min Scan# 3313
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:146 Resp: 34483
Ion Ratio Lower Upper
146 100
111 45.8 32.6 48.8
148 62.2 51.9 77.9









#83
1,4-Dichlorobenzene
Concen: 1.779 ug/l
RT: 11.832 min Scan# 31
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:146 Resp: 34219
Ion Ratio Lower Upper
146 100
111 45.9 29.5 44.3#
148 62.8 51.4 77.0



#84
n-Butylbenzene
Concen: 2.329 ug/l
RT: 12.202 min Scan# 3456
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 91 Resp: 71416
Ion Ratio Lower Upper
91 100
92 53.0 43.4 65.0
134 23.2 21.2 31.8









#85
1,2-Dichlorobenzene
Concen: 1.746 ug/l
RT: 12.209 min Scan# 341
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:146 Resp: 31571
Ion Ratio Lower Upper
146 100
111 49.4 20.8 62.5
148 65.7 31.4 94.2



#86
1,2-Dibromo-3-Chloropropane
Concen: 2.109 ug/l
RT: 12.990 min Scan# 3701
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion: 75 Resp: 3136
Ion Ratio Lower Upper
75 100
155 65.8 66.0 99.0#
157 85.6 85.9 128.9#









#87
1,2,4-Trichlorobenzene
Concen: 1.906 ug/l
RT: 13.836 min Scan# 3964
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26




Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01


Tgt Ion:	180	Resp:	20672
Ion Ratio	Lower	Upper	
180	100		
182	97.5	74.6	112.0
145	32.9	23.4	35.0



#88
Hexachlorobutadiene
Concen: 1.441 ug/l
RT: 14.016 min Scan# 4020
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:	225	Resp:	10953
Ion Ratio	Lower	Upper	
225	100		
223	63.7	49.8	74.6
227	64.6	51.5	77.3









#89
Naphthalene
Concen: 2.174 ug/l
RT: 14.083 min Scan# 4011
Delta R.T. 0.001 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Instrument :
MSVOA_U
ClientSampleId :
VU0209WBS01

Tgt Ion:128 Resp: 45708
Ion Ratio Lower Upper
128 100
127 12.9 10.2 15.4
129 10.3 8.6 12.8



#90
1,2,3-Trichlorobenzene
Concen: 1.870 ug/l
RT: 14.324 min Scan# 4116
Delta R.T. -0.003 min
Lab File: VU053015.D
Acq: 09 Feb 2023 13:26

Tgt Ion:180 Resp: 19249
Ion Ratio Lower Upper
180 100
182 95.4 77.8 116.6
145 34.9 25.2 37.8

