

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
 Data File : VU053025.D
 Acq On : 09 Feb 2023 18:53
 Operator : JC/MD
 Sample : 01479-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L10-U-020723

Quant Time: Feb 10 01:23:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.112	96	31243	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.630	95	11151	1.073	ug/l	0.00
Spiked Amount 1.000			Recovery	=	107.000%	
68) 1,2-Dichlorobenzene-d4	12.189	152	9360	0.876	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.585	101	2192	0.284	ug/l #	87
15) Methylene Chloride	3.042	84	2021	0.230	ug/l	90
22) cis-1,2-Dichloroethene	4.659	96	1864	0.220	ug/l	97
25) Methyl tert-Butyl Ether	3.363	73	4428	0.265	ug/l #	85
37) Trichloroethene	6.540	130	8942	0.989	ug/l	85
58) Tetrachloroethene	8.550	164	7771	0.952	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
Data File : VU053025.D
Acq On : 09 Feb 2023 18:53
Operator : JC/MD
Sample : 01479-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L10-U-020723

Quant Time: Feb 10 01:23:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration

