

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
 Data File : VU053021.D
 Acq On : 09 Feb 2023 16:24
 Operator : JC/MD
 Sample : 01479-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-020723

Quant Time: Feb 10 01:22:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.109	96	32725	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.630	95	11250	1.033	ug/l	0.00
Spiked Amount 1.000			Recovery	=	103.000%	
68) 1,2-Dichlorobenzene-d4	12.189	152	9750	0.871	ug/l	0.00
Spiked Amount 1.000			Recovery	=	87.000%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020923\
Data File : VU053021.D
Acq On : 09 Feb 2023 16:24
Operator : JC/MD
Sample : 01479-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-020723

Quant Time: Feb 10 01:22:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.109 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU053021.D
Acq: 09 Feb 2023 16:24




Instrument :
MSVOA_U
ClientSampleId :
TB-01-020723


Tgt Ion: 96	Resp: 32725
Ion Ratio	Lower Upper
96 100	
70 21.6	16.2 24.4
95 9.9	7.4 11.0
97 6.5	0.0 0.0



#57
4-Bromofluorobenzene
Concen: 1.033 ug/l
RT: 10.630 min Scan# 2967
Delta R.T. 0.001 min
Lab File: VU053021.D
Acq: 09 Feb 2023 16:24

Tgt Ion: 95	Resp: 11250
Ion Ratio	Lower Upper
95 100	
174 63.0	66.9 100.3
176 63.1	61.9 92.9





#68

1,2-Dichlorobenzene-d4

Concen: 0.871 ug/l

RT: 12.189 min Scan# 34

Delta R.T. 0.001 min

Lab File: VU053021.D


Acq: 09 Feb 2023 16:24

Instrument :

MSVOA_U

ClientSampleId :

TB-01-020723



Tgt Ion:152 Resp: 9750

Ion Ratio Lower Upper

152 100

115 70.1 0.0 262.6

150 155.9 0.0 638.0

