

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021119\
 Data File : VU029400.D
 Acq On : 11 Feb 2019 16:45
 Operator : JC/SP
 Sample : K1436-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESX94

Quant Time: Feb 13 11:34:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Feb 12 01:17:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	238739	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	225093	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	119578	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	82654	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.68	69	68225	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.27	63	127042	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	4.19	46	157953	49.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.54%
24) Chloroform-d	4.65	84	128906	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
26) 1,2-Dichloroethane-d4	5.31	65	64951	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.34	84	270834	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.33	67	79719	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.56	98	264855	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	7.85	79	35119	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
46) 2-Hexanone-d5	8.31	63	161958	50.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.36%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	65830	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	112440	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds					Ovalue
13) Acetone	2.32	43	6358	2.504 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021119\
Data File : VU029400.D
Acq On : 11 Feb 2019 16:45
Operator : JC/SP
Sample : K1436-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESX94

Quant Time: Feb 13 11:34:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Feb 12 01:17:05 2019
Response via : Initial Calibration

