

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021119\
 Data File : VU029401.D
 Acq On : 11 Feb 2019 17:09
 Operator : JC/SP
 Sample : K1436-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESX95

Quant Time: Feb 13 11:34:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Feb 12 01:17:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	237871	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	224958	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	118666	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	72719	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.68	69	62453	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.27	63	118008	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	4.19	46	148148	46.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.70%
24) Chloroform-d	4.65	84	117303	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.31	65	60255	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
32) Benzene-d6	5.34	84	247801	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.33	67	72686	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	7.56	98	241084	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	7.85	79	31821	4.22	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.40%
46) 2-Hexanone-d5	8.31	63	153436	47.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.14%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	59615	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	102938	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%


Target Compounds					Ovalue
13) Acetone	2.32	43	4699	1.857 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021119\
Data File : VU029401.D
Acq On : 11 Feb 2019 17:09
Operator : JC/SP
Sample : K1436-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESX95

Quant Time: Feb 13 11:34:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Feb 12 01:17:05 2019
Response via : Initial Calibration





#13

Acetone

Concen: 1.857 ug/L

RT: 2.32 min Scan# 384

Delta R.T. 0.00 min

Lab File: VU029401.D

Acq: 11 Feb 2019 17:09

Instrument :
MSVOA_U
ClientSampleId :
ESX95

Tot Ion: 43 Resp: 4699

Ion Ratio Lower Upper

43 100

58 28.6 0.0 64.2

