

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021220\
 Data File : VU036737.D
 Acq On : 11 Feb 2020 22:04
 Operator : JC/MD
 Sample : L1490-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B75

Quant Time: Feb 12 08:51:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Feb 12 08:13:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	118790	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	110771	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	46098	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	53211	5.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	1.93	69	46156	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.00%
11) 1,1-Dichloroethene-d2	2.59	63	67120	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	4.69	46	118373	55.25	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.50%
24) Chloroform-d	5.11	84	78780	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.75	65	46301	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.77	84	173205	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
36) 1,2-Dichloropropane-d6	6.73	67	54795	5.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.60%
41) Toluene-d8	7.93	98	158147	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
43) trans-1,3-Dichloropropene-	8.21	79	19524	4.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.00%
46) 2-Hexanone-d5	8.67	63	98729	52.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.08%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	41524	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	12.21	152	49958	5.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.40%

Target Compounds				Ovalue
3) Chloromethane	1.54	50	3284	0.402 ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021220\
Data File : VU036737.D
Acq On : 11 Feb 2020 22:04
Operator : JC/MD
Sample : L1490-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B75

Quant Time: Feb 12 08:51:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Feb 12 08:13:11 2020
Response via : Initial Calibration

