

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021220\
 Data File : VU036741.D
 Acq On : 12 Feb 2020 10:47
 Operator : JC/MD
 Sample : VU0212WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK13

Quant Time: Feb 13 01:14:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 13 00:59:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	118580	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	111332	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	45851	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	51059	5.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.20%
7) Chloroethane-d5	1.93	69	44758	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.59	63	66585	3.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.20%
20) 2-Butanone-d5	4.69	46	116260	54.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.72%
24) Chloroform-d	5.11	84	77854	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.75	65	44454	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.77	84	174357	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.00%
36) 1,2-Dichloropropane-d6	6.73	67	55599	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.93	98	156970	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
43) trans-1,3-Dichloropropene-	8.21	79	20779	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	8.67	63	94972	49.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.60%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	40805	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	47831	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021220\
Data File : VU036741.D
Acq On : 12 Feb 2020 10:47
Operator : JC/MD
Sample : VU0212WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK13

Quant Time: Feb 13 01:14:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 13 00:59:33 2020
Response via : Initial Calibration

