

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU021419\
 Data File : VU029471.D
 Acq On : 14 Feb 2019 21:30
 Operator : JC/SP
 Sample : K1438-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESXC3

Quant Time: Feb 15 03:10:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 15 01:55:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	188661	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	175294	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	89933	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	47652	3.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.20%
7) Chloroethane-d5	1.68	69	47249	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.40%
11) 1,1-Dichloroethene-d2	2.27	63	79714	2.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.80%#
20) 2-Butanone-d5	4.19	46	130656	49.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.16%
24) Chloroform-d	4.65	84	92163	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.31	65	48224	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
32) Benzene-d6	5.34	84	192006	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.33	67	58736	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.56	98	180552	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	7.85	79	22004	3.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.20%
46) 2-Hexanone-d5	8.31	63	135711	52.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.96%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	51232	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	80500	4.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU021419\
Data File : VU029471.D
Acq On : 14 Feb 2019 21:30
Operator : JC/SP
Sample : K1438-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESXC3

Quant Time: Feb 15 03:10:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 15 01:55:09 2019
Response via : Initial Calibration

