

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021420\
 Data File : VU036790.D
 Acq On : 14 Feb 2020 16:31
 Operator : JC/MD
 Sample : L1537-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AQ3

Quant Time: Feb 15 00:55:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 14 23:39:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	123221	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	114182	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	45007	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	40897	3.86	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.20%
7) Chloroethane-d5	1.93	69	40250	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.60%
11) 1,1-Dichloroethene-d2	2.59	63	58031	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	4.67	46	116250	52.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.62%
24) Chloroform-d	5.11	84	75524	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.74	65	45025	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.76	84	165003	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.73	67	54038	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	7.93	98	145790	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	8.21	79	19082	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	8.66	63	98240	50.23	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.46%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	41379	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	44537	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021420\
Data File : VU036790.D
Acq On : 14 Feb 2020 16:31
Operator : JC/MD
Sample : L1537-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AQ3

Quant Time: Feb 15 00:55:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 14 23:39:59 2020
Response via : Initial Calibration

