

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021420\  
 Data File : VU036780.D  
 Acq On : 14 Feb 2020 10:45  
 Operator : JC/MD  
 Sample : VU0214WBL01  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK15

Quant Time: Feb 14 23:49:26 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR021220WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Feb 14 23:39:59 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	127174	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	116426	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	48607	5.00	ug/L	0.00

## System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	46280	4.23	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.60%
7) Chloroethane-d5	1.93	69	42960	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	2.59	63	64170	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	4.69	46	131507	57.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.68%
24) Chloroform-d	5.11	84	81936	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	5.74	65	47477	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.76	84	175741	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.73	67	57248	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.92	98	160250	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
43) trans-1,3-Dichloropropene-	8.21	79	20489	4.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.80%
46) 2-Hexanone-d5	8.67	63	108452	54.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.76%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	43461	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
64) 1,2-Dichlorobenzene-d4	12.21	152	49342	5.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.60%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021420\  
 Data File : VU036780.D  
 Acq On : 14 Feb 2020 10:45  
 Operator : JC/MD  
 Sample : VU0214WBL01  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampled :  
 VBLK15

Quant Time: Feb 14 23:49:26 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR021220WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Feb 14 23:39:59 2020  
 Response via : Initial Calibration

