

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021420\
 Data File : VU036792.D
 Acq On : 14 Feb 2020 17:20
 Operator : JC/MD
 Sample : L1537-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AQ5

Quant Time: Feb 15 00:46:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 14 23:39:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	116792	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	107757	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	42473	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	42737	4.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.00%
7) Chloroethane-d5	1.93	69	40876	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.59	63	58734	3.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.00%
20) 2-Butanone-d5	4.67	46	117475	55.77	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.54%
24) Chloroform-d	5.11	84	75776	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.74	65	43705	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.76	84	164174	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	6.73	67	53704	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.92	98	147127	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	8.21	79	18708	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
46) 2-Hexanone-d5	8.66	63	97416	52.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.56%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	42013	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	45327	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021420\
Data File : VU036792.D
Acq On : 14 Feb 2020 17:20
Operator : JC/MD
Sample : L1537-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AQ5

Quant Time: Feb 15 00:46:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 14 23:39:59 2020
Response via : Initial Calibration

