

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
 Data File : VU047136.D
 Acq On : 14 Feb 2022 16:58
 Operator : SY/MD
 Sample : N1205-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Feb 17 14:31:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Feb 17 14:31:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	55794	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	20398	0.944	ug/l	0.00
Spiked Amount 1.000			Recovery	=	94.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	20901	0.929	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	344423	26.879	ug/l	94
24) tert-Butyl Alcohol	3.253	59	44256	62.999	ug/l #	77
25) Methyl tert-Butyl Ether	3.369	73	1546063	41.805	ug/l	100
31) Isopropyl Ether	4.002	45	290126	7.113	ug/l	96
32) Ethyl-t-butyl ether	4.510	59	1283070	30.946	ug/l	98
33) Tert-Amyl methyl ether	5.944	73	1746646	45.737	ug/l	99
73) n-propylbenzene	10.909	120	887754	44.693	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
Data File : VU047136.D
Acq On : 14 Feb 2022 16:58
Operator : SY/MD
Sample : N1205-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Feb 17 14:31:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Feb 17 14:31:24 2022
Response via : Initial Calibration

